

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 09 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/9/1	4.1	4.9	3.7	4.5	3.9	4.9	-19.7	-16.8	NA	NA	18.8	19.8
2020/9/2	4.2	4.9	3.6	5.1	3.6	4.7	-21	-18.5	NA	NA	18.8	19.8
2020/9/3	4.1	5.6	3.6	4.5	3.6	4.6	-21.2	-18.7	NA	NA	18.8	19.7
2020/9/4	3.9	4.7	3.6	4.6	3.9	4.7	-19.7	-18.7	NA	NA	18.8	19.5
2020/9/5	4.3	4.7	3.9	4.6	4	4.6	-19.4	-19.3	NA	NA	18.7	18.9
2020/9/6	4.3	4.7	3.9	4.7	4	4.7	-19.4	-19.3	NA	NA	18.4	18.7
2020/9/7	3.9	4.7	3.6	4.7	3.9	4.7	-20.7	-18.6	NA	NA	18.3	19.3
2020/9/8	4	4.6	3.5	4.6	3.8	4.7	-21.2	-19.7	NA	NA	18.4	19.8
2020/9/9	3.9	4.7	3.6	4.7	3.8	4.8	-21.7	-19.9	NA	NA	18.4	19.3
2020/9/10	4	4.7	3.6	4.7	3.6	4.8	-20.3	-19.6	NA	NA	18.6	20.3
2020/9/11	4.2	4.8	3.4	4.7	3.8	5	-19.7	-19.4	NA	NA	19.4	20.2
2020/9/12	4.2	4.7	4	4.6	3.9	4.7	-19.7	-19.6	NA	NA	19.4	19.6
2020/9/13	4.2	4.6	4	4.6	3.9	4.7	-19.7	-19.6	NA	NA	19.4	19.4
2020/9/14	4.3	4.7	3.7	5.1	3.6	4.7	-20.8	-19.6	NA	NA	19.3	20.1
2020/9/15	4.3	4.7	3.7	4.5	3.4	4.6	-21.1	-20.6	NA	NA	19.4	20.5
2020/9/16	4.2	4.7	3.7	4.5	3.3	4.7	-21.2	-20.8	NA	NA	18.9	19.8
2020/9/17	4.3	4.7	3.7	4.6	3.6	4.7	-20.9	-19.9	NA	NA	19.2	20.1
2020/9/18	4.2	4.7	3.8	4.5	3.5	4.7	-20.9	-19.5	NA	NA	19.3	20.1
2020/9/19	4.3	4.7	3.9	4.4	3.9	4.7	-20.4	-19.9	NA	NA	19.3	19.6
2020/9/20	4.3	4.7	3.9	4.4	3.9	4.7	-20.4	-20.1	NA	NA	19.2	19.3
2020/9/21	4.3	4.8	3.7	4.6	3.6	4.7	-20.4	-19.6	NA	NA	19.1	19.8
2020/9/22	4.3	4.9	3.7	4.5	3.7	4.7	-20.2	-19.5	NA	NA	19.3	19.9
2020/9/23	4.4	4.9	3.7	4.8	3.4	5.3	-20.2	-19.3	NA	NA	19.4	20
2020/9/24	3.9	5.5	3.7	4.3	3.5	4.5	-19.6	-18.8	NA	NA	19.4	20.2
2020/9/25	3.8	4.6	3.7	4.3	3.8	4.8	-19.9	-19	NA	NA	19.6	20.1
2020/9/26	4.2	4.6	3.9	4.4	3.9	4.5	-20.1	-19.8	NA	NA	19.5	19.9
2020/9/27	4.2	4.6	3.9	4.3	3.9	4.5	-20.2	-19.9	NA	NA	19.3	19.6
2020/9/28	3.9	4.6	3.7	4.5	3.4	5.5	-20.2	-18.5	NA	NA	19.2	19.9
2020/9/29	4.1	4.6	3.8	4.4	3.2	4.9	-20.4	-18.4	NA	NA	19.4	21.6
2020/9/30	3.9	4.7	3.8	4.4	3.3	4.7	-20.2	-19.6	NA	NA	19.6	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部邱保祥 1005
師(三)級藥師 4100 藥學部莊承芳 1005
師(三)級藥師 1358

溫度均在正常範圍內