

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 08 月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2020/8/1	4.2	4.8	4.5	5.4	21.4	21.6
2020/8/2	4.2	4.8	4.5	5.4	21.2	21.4
2020/8/3	4.2	4.8	4.8	5.7	21.3	21.7
2020/8/4	4.2	4.8	4.9	5.9	21.6	21.9
2020/8/5	4.2	4.8	4.9	6	21.8	22
2020/8/6	4.3	4.9	4.8	5.8	21.3	21.8
2020/8/7	4.4	5.2	4.9	5.8	21.1	21.3
2020/8/8	4.3	4.9	4.9	5.7	21.1	21.2
2020/8/9	4.4	4.9	4.9	5.7	20.9	21.1
2020/8/10	4.4	5.1	4.6	5.8	20.8	21.1
2020/8/11	4.5	5.1	4.6	5.6	20.9	21.3
2020/8/12	4.5	5.2	4.5	5.6	21.1	21.2
2020/8/13	4.3	5.2	4.6	5.8	21.1	21.2
2020/8/14	4.3	5.2	4.5	5.9	21.1	21.4
2020/8/15	4.3	4.8	4.5	5.4	21.1	21.3
2020/8/16	4.2	4.8	4.5	5.5	20.7	21.1
2020/8/17	4.3	4.9	4.5	5.7	20.6	20.9
2020/8/18	4	5.6	4.5	5.6	20.9	21.3
2020/8/19	3.9	5.2	4.2	5.5	21.2	21.5
2020/8/20	3.8	4.7	3.9	5.4	21.3	21.6
2020/8/21	3.6	4.6	4.4	5.6	21.4	21.7
2020/8/22	3.6	4.4	4.4	5.4	21.4	21.7
2020/8/23	3.7	4.4	4.4	5.4	21.3	21.4
2020/8/24	3.7	4.5	4.2	5.6	21.3	21.4
2020/8/25	3.6	4.4	4.3	5.4	21.7	22.4
2020/8/26	3.6	4.4	4.6	5.7	21.8	22
2020/8/27	3.6	4.8	4.6	5.6	21.8	21.9
2020/8/28	3.7	4.4	4.7	5.6	21.5	21.8
2020/8/29	3.7	4.4	4.7	5.6	21.1	21.5
2020/8/30	3.7	4.4	4.7	5.6	20.7	21.1
2020/8/31	3.8	4.6	4.7	5.6	20.6	20.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥
師(三級藥師) 109
16242 藥學部 溫雅如
師(三級藥師) 109
0901
1616