

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 07 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/7/1	4.2	4.8	3.6	4.5	4.7	5.3	-19.9	-18.1	NA	NA	19.6	20.2
2020/7/2	4.3	4.9	3.7	4.5	4.7	5.3	-19.9	-19.4	NA	NA	19.4	20.1
2020/7/3	4.3	4.9	3.7	4.5	4.6	5.3	-19.9	-19.3	NA	NA	19.2	20.1
2020/7/4	4.4	4.6	3.7	4.5	4.8	5.2	-19.7	-19.6	NA	NA	19.2	19.4
2020/7/5	4.4	4.6	3.7	4.5	4.8	5.2	-19.7	-19.6	NA	NA	19.1	19.2
2020/7/6	4.3	4.8	3.6	4.5	4.4	5.2	-19.7	-19.1	NA	NA	19	20.1
2020/7/7	4.3	4.9	3.7	4.4	4.6	5.2	-20.3	-19.4	NA	NA	19.1	20.1
2020/7/8	4.3	4.8	3.7	4.4	4.6	5.2	-20.3	-19.8	NA	NA	19.2	20.2
2020/7/9	4.2	4.8	3.7	4.5	4.3	5.2	-20.2	-19.8	NA	NA	19.2	20.2
2020/7/10	4.2	4.7	3.7	4.4	4.3	4.8	-20.3	-19.8	NA	NA	19.2	20.1
2020/7/11	4.3	4.7	3.7	4.4	4.3	4.8	-20.3	-20	NA	NA	19.2	19.4
2020/7/12	4.3	4.7	3.7	4.4	4.3	4.8	-20.3	-20	NA	NA	19	19.2
2020/7/13	4.1	4.7	3.7	4.5	4.2	4.9	-20.3	-19.4	NA	NA	18.9	20
2020/7/14	4.1	4.7	3.7	4.5	4.3	4.9	-20.3	-19.6	NA	NA	19.2	19.7
2020/7/15	4.2	4.8	3.7	4.5	4.3	4.8	-20.4	-19.6	NA	NA	19	19.8
2020/7/16	4.2	4.8	3.7	4.5	4.3	5.1	-20.4	-20.1	NA	NA	19.2	19.9
2020/7/17	4	4.9	3.7	4.5	4.4	5.1	-20.4	-19.6	NA	NA	18.9	19.7
2020/7/18	4.3	4.7	3.9	4.5	4.5	5.1	-19.7	-19.6	NA	NA	18.9	19.1
2020/7/19	4.3	4.7	3.9	4.5	4.6	5.1	-19.8	-19.6	NA	NA	18.8	18.8
2020/7/20	4	4.7	3.7	4.5	4.4	5.2	-20.4	-19.7	NA	NA	17.9	19.3
2020/7/21	4	4.7	3.8	4.6	4.4	5	-20.8	-20.2	NA	NA	17.7	18.3
2020/7/22	3.9	4.8	3.7	4.5	2.6	5.1	-20.8	-19.8	NA	NA	17.5	18.3
2020/7/23	4.1	4.8	3.8	4.7	4.3	5.1	-20.1	-19.6	NA	NA	17.4	19.9
2020/7/24	4.1	4.7	3.8	4.6	4.4	5.8	-20.1	-19.3	NA	NA	19.3	20.4
2020/7/25	4.4	4.7	3.9	4.5	4.4	4.9	-19.9	-19.7	NA	NA	19.3	19.8
2020/7/26	4.3	4.7	3.9	4.5	4.4	4.8	-19.9	-19.7	NA	NA	19.1	19.3
2020/7/27	4.1	4.8	3.7	4.6	3.9	5.9	-19.9	-19.6	NA	NA	18.9	19.8
2020/7/28	4.2	4.8	3.7	4.7	3.9	4.9	-20.1	-19.6	NA	NA	19	20.1
2020/7/29	4.2	4.8	3.7	4.6	3.9	4.9	-20.4	-19.1	NA	NA	19.2	20.1
2020/7/30	4.2	4.9	3.7	4.6	4	4.9	-20.6	-19.9	NA	NA	19.2	20.1
2020/7/31	4.2	4.9	3.8	4.7	4	4.9	-20.7	-18.3	NA	NA	19.3	20.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥

師(三)級藥師

10908030900

藥學部 溫雅如

師(三)級藥師

1090803438