

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 06 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/6/1	4.1	4.9	3.4	4.7	4.4	5.1	-20.8	-19.7	NA	NA	19	19.9
2020/6/2	4.2	4.9	3.4	4.9	4.4	5.1	-20.6	-20.1	NA	NA	19.2	20.1
2020/6/3	4.2	4.9	3.4	4.6	4.3	5.1	-20.6	-20.2	NA	NA	19.3	19.9
2020/6/4	4.2	4.9	3.4	4.6	4.4	5.1	-20.6	-20	NA	NA	19.1	19.8
2020/6/5	4	5.1	3.4	4.6	4.3	4.9	-20.4	-19.8	NA	NA	19.1	19.8
2020/6/6	4.4	4.8	4.1	4.6	4.4	4.9	-20.3	-19.9	NA	NA	19	19.2
2020/6/7	4.4	4.8	3.9	4.6	4.4	4.9	-20.4	-19.9	NA	NA	18.9	19.3
2020/6/8	4	4.8	3.3	4.7	4.3	5	-20.3	-19.8	NA	NA	19.3	20.2
2020/6/9	4.1	4.8	3.5	4.7	4.4	4.9	-20.6	-19.6	NA	NA	19.5	20.4
2020/6/10	4.2	4.9	3.6	4.6	4.4	4.9	-20.6	-20.1	NA	NA	19.7	20.4
2020/6/11	4.3	5.1	3.4	4.7	4.3	4.9	-20.6	-18.8	NA	NA	18.9	20.3
2020/6/12	4.3	5.1	3.5	4.6	4.4	4.9	-20.4	-19.8	NA	NA	18.7	19.2
2020/6/13	4.7	5.1	3.9	4.6	4.6	4.9	-20.6	-20.2	NA	NA	18.4	18.8
2020/6/14	4.7	5.1	3.9	4.6	4.4	4.9	-20.6	-20.2	NA	NA	18.2	18.4
2020/6/15	4.2	5.1	3.4	4.6	4.4	5.1	-20.6	-19.8	NA	NA	18.2	19.3
2020/6/16	4	5	3.4	4.5	4.4	5.1	-23.9	-19.6	NA	NA	18.4	19.6
2020/6/17	4.1	4.9	3.3	4.6	4.4	5.1	-23.7	-21.9	NA	NA	18.6	19.3
2020/6/18	4	4.7	3.3	4.3	4.4	5.2	-22.4	-21.3	NA	NA	18.6	19.3
2020/6/19	4.1	4.7	3.3	4.3	4.4	5.2	-22.7	-21.4	NA	NA	18.6	19.3
2020/6/20	4.4	4.7	3.5	4.3	4.7	5.3	-22.7	-22.2	NA	NA	18.7	19.1
2020/6/21	4.4	4.7	3.6	4.3	4.7	5.2	-22.7	-22.2	NA	NA	18.6	18.7
2020/6/22	4.1	4.7	3.3	4.3	4.5	5.2	-22.7	-21.9	NA	NA	18.4	19.3
2020/6/23	4.1	4.8	3.2	4.3	4.5	5.2	-22.4	-21.2	NA	NA	18.7	19.4
2020/6/24	4	4.8	3.3	4.3	4.5	5.2	-22.5	-21.4	NA	NA	18.4	19.2
2020/6/25	4.4	4.8	3.5	4.3	4.8	5.2	-22.4	-21.9	NA	NA	18.1	18.4
2020/6/26	4.4	4.8	3.6	4.3	4.8	5.2	-22.5	-21.9	NA	NA	17.9	18.1
2020/6/27	4.4	4.8	3.6	4.3	4.7	5.2	-22.5	-22	NA	NA	17.8	17.9
2020/6/28	4.4	4.8	3.5	4.3	4.7	5.3	-22.5	-21.9	NA	NA	17.7	18.8
2020/6/29	4.1	4.8	3.3	4.5	4.7	5.3	-22.4	-19.4	NA	NA	18.8	19.9
2020/6/30	4.2	4.9	3.4	4.4	4.7	5.3	-19.9	-19.2	NA	NA	19.2	20.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 10907010436
藥學部 溫雅如 10907010436