

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 06 月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2020/6/1	4	4.9	3.5	5.7	21.1	21.4
2020/6/2	3.9	4.9	4.6	5.8	21.3	21.7
2020/6/3	4	4.7	4.7	6.1	21.6	21.7
2020/6/4	3.9	4.7	4.9	6	21.4	21.7
2020/6/5	3.9	4.7	4.8	6	21.5	21.7
2020/6/6	3.9	4.6	4.8	5.8	21.4	21.7
2020/6/7	3.9	4.6	4.8	5.8	21.3	21.4
2020/6/8	3.8	4.6	4.8	6.1	21.2	21.7
2020/6/9	3.9	4.5	4.8	5.9	21.5	21.8
2020/6/10	3.9	4.6	4.9	5.9	21.7	21.9
2020/6/11	3.9	4.7	4.8	5.9	21.7	21.9
2020/6/12	4	4.7	4.8	5.7	21.3	21.7
2020/6/13	4	4.7	4.9	5.6	21.1	21.4
2020/6/14	4	4.7	4.9	5.7	20.7	21.1
2020/6/15	4.1	4.8	4.7	5.9	20.4	20.9
2020/6/16	3.9	4.9	4.7	5.7	20.7	21.1
2020/6/17	4.1	4.9	4.4	5.9	20.9	21.2
2020/6/18	4.1	4.8	4.7	5.8	21.1	21.3
2020/6/19	4.1	4.8	4.7	5.9	21.1	21.3
2020/6/20	4	4.8	4.7	5.8	21.1	21.3
2020/6/21	4	4.7	4.7	5.8	20.9	21.2
2020/6/22	4.1	4.7	4.6	5.9	20.7	20.9
2020/6/23	4	4.8	4.6	5.8	20.8	21.1
2020/6/24	4	4.9	4.6	5.8	21.1	21.2
2020/6/25	4.2	4.8	4.7	5.8	20.8	21.1
2020/6/26	4.2	4.8	4.7	5.8	20.3	20.7
2020/6/27	4.2	4.9	4.7	5.9	20.1	20.3
2020/6/28	4.2	4.8	4.7	5.9	20.1	20.4
2020/6/29	4.2	4.9	4.6	5.9	20.4	20.9
2020/6/30	4.2	4.9	4.6	5.8	20.9	21.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥
師(三)級藥師 1090920
藥學部 溫雅如
師(三)級藥師 1090524