

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 109 年 06 月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2020/6/1	4.4	5	19.4	19.9
2020/6/2	4.4	5	19.5	20.1
2020/6/3	4.4	5.1	19.7	20
2020/6/4	4.3	4.9	19.7	20.1
2020/6/5	4.7	5.1	19.8	20.3
2020/6/6	4.7	5	19.7	20
2020/6/7	4.7	5	19.7	19.9
2020/6/8	4.4	5.1	19.7	20.1
2020/6/9	4.4	4.9	19.8	20.1
2020/6/10	4.6	5	19.7	20.1
2020/6/11	4.5	5.1	19.6	20.3
2020/6/12	4.7	5.2	19.8	20.4
2020/6/13	4.8	5.2	19.7	20.1
2020/6/14	4.9	5.2	19.6	19.8
2020/6/15	4.8	5.3	19.6	20.2
2020/6/16	4.7	5.3	19.7	20.1
2020/6/17	4.7	5.2	19.7	20.3
2020/6/18	4.7	5.3	20.1	20.4
2020/6/19	4.7	5.3	20.1	20.6
2020/6/20	4.7	5.2	20.1	20.7
2020/6/21	4.7	5.2	19.9	20.1
2020/6/22	4.7	5.2	19.8	20.9
2020/6/23	4.7	5.2	19.9	21.1
2020/6/24	4.7	5.2	20.2	21.2
2020/6/25	4.7	5.2	20.1	20.4
2020/6/26	4.7	5.1	19.9	20.4
2020/6/27	4.7	5.1	19.8	20.2
2020/6/28	4.7	5.2	19.7	20.2
2020/6/29	4.7	5.2	19.7	20.7
2020/6/30	4.7	5.1	19.9	21.1

註1：溫度記錄器型號：ESCORT/ Model No.: EI-IN-D-32-L

註2：溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師：**藥學部邱保祥**
109
07/01
(三)級藥師

藥學部溫雅如
109
07/01
(二)級藥師