

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109年05月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/5/1	4.4	5.4	4.2	5.1	4.3	5.3	-20.7	-19	NA	NA	19.1	19.8
2020/5/2	5.1	5.4	4.5	5	4.3	4.8	-20.8	-20.3	NA	NA	18.8	19.1
2020/5/3	5.1	5.4	4.5	5.1	4.3	4.8	-20.7	-20.4	NA	NA	18.8	18.8
2020/5/4	4.4	5.3	4.2	5.1	4.3	5.2	-20.7	-19.7	NA	NA	18.7	19.6
2020/5/5	4.4	5.3	4.2	5.3	4.3	5.3	-20.3	-18.8	NA	NA	18.4	19.2
2020/5/6	4.4	5.3	4.3	5.2	4.3	5.3	-19.6	-18.6	NA	NA	18.2	21.2
2020/5/7	4.4	5.3	4.2	5.1	4.3	5.3	-21.2	-19.4	NA	NA	18	18.7
2020/5/8	4.4	5.3	4.2	5.1	4.3	5.3	-21.4	-20	NA	NA	17.9	18.6
2020/5/9	4.9	5.2	4.7	5.1	4.3	4.8	-21.5	-21.2	NA	NA	17.7	18
2020/5/10	4.8	5.2	4.7	5.1	4.3	4.8	-21.4	-21.2	NA	NA	17.5	17.7
2020/5/11	4.3	5.2	4.2	5.2	4.3	5.2	-21.5	-19.7	NA	NA	17.4	18.2
2020/5/12	4.4	5.3	4.2	5.1	4.3	5.3	-20.8	-20.4	NA	NA	17.6	19.2
2020/5/13	4.4	5.3	4.2	5.2	4.3	5.2	-20.8	-20.4	NA	NA	18.7	19.4
2020/5/14	4.3	5.3	4.2	5.4	4.3	5.2	-20.8	-20.2	NA	NA	18.8	19.6
2020/5/15	4.3	5.2	4.1	5.2	4.3	5.1	-20.6	-20.2	NA	NA	18.9	19.6
2020/5/16	4.9	5.3	4.5	5.1	4.4	4.9	-20.4	-20.2	NA	NA	18.8	19.1
2020/5/17	4.9	5.2	4.4	5.1	4.4	4.9	-20.4	-20.2	NA	NA	18.7	18.8
2020/5/18	4.3	5.3	4.2	5	4.4	5.1	-20.6	-18.4	NA	NA	18.7	19.6
2020/5/19	4.4	5.3	4.2	5.1	4.2	5.1	-20.7	-20.2	NA	NA	18.8	19.7
2020/5/20	4.3	5.3	4.2	5.1	4.2	5	-20.7	-20.2	NA	NA	18.9	19.7
2020/5/21	4.3	5.4	4.2	5.1	4.3	4.9	-20.4	-19.9	NA	NA	18.9	19.7
2020/5/22	4.5	5.5	3.2	5.2	4.3	4.8	-20.4	-19.6	NA	NA	19	19.8
2020/5/23	5.1	5.4	3.8	4.5	4.4	4.8	-20.4	-20	NA	NA	19	19.2
2020/5/24	5.1	5.4	3.7	4.5	4.4	4.8	-20.4	-20.1	NA	NA	18.8	19
2020/5/25	4.5	5.4	3.1	4.5	4.4	5.3	-20.4	-20	NA	NA	18.7	20.3
2020/5/26	4.7	5.4	3.1	4.4	4.6	5.3	-20.9	-20.2	NA	NA	18.9	19.7
2020/5/27	4.4	5.6	3	4.4	4.4	5.3	-20.9	-20.4	NA	NA	19	19.7
2020/5/28	4.2	5.2	3.4	4.7	4.5	5.2	-20.9	-20.4	NA	NA	19	19.9
2020/5/29	4.1	5.1	3.4	4.7	4.5	5.1	-20.9	-20.6	NA	NA	19.1	19.8
2020/5/30	4.7	5.1	4	4.7	4.7	5.1	-20.9	-20.6	NA	NA	19.1	19.2
2020/5/31	4.7	5.2	4	4.7	4.7	5.1	-20.9	-20.6	NA	NA	19	19.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 10906010905
藥學部 四時如 109