

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109年05月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2020/5/1	4.8	5.5	3	4.4	20.9	21.1
2020/5/2	4.8	5.5	3.2	4.3	20.9	21.1
2020/5/3	4.8	5.5	3.2	4.3	20.8	20.9
2020/5/4	4.8	5.6	3.2	4.3	20.7	21.1
2020/5/5	4.8	5.6	3	4.3	20.9	21.2
2020/5/6	4.7	5.6	3.1	5.5	20.9	21.1
2020/5/7	4.7	5.4	3.1	4.3	20.8	21.1
2020/5/8	4.4	5.4	3.2	4.2	20.7	20.9
2020/5/9	4.4	5.2	3.3	4.3	20.5	20.8
2020/5/10	4.4	5.2	3.4	4.3	20.2	20.5
2020/5/11	4.4	5.3	3.2	4.4	20.1	20.3
2020/5/12	4.4	5.3	3.1	4.2	20.2	20.7
2020/5/13	4.4	5.3	3.1	4.1	20.7	21.1
2020/5/14	4.4	5.3	2.9	4.2	20.9	21.3
2020/5/15	4.4	5.2	2.9	4.2	21.2	21.5
2020/5/16	4.4	5.2	3.1	4.2	21.3	21.4
2020/5/17	4.4	5.2	3.2	4.2	21	21.3
2020/5/18	4.4	5.2	3.3	4.5	20.9	21.2
2020/5/19	4.4	5.2	3.3	4.4	21.1	21.3
2020/5/20	4.4	5.3	3.3	4.4	21.2	21.4
2020/5/21	4.4	5.3	3.3	4.4	21.2	21.4
2020/5/22	4.4	5.3	3.3	4.4	21.3	21.5
2020/5/23	4.5	5.3	3.6	4.4	21.3	21.5
2020/5/24	4.5	5.3	3.6	4.4	20.9	21.3
2020/5/25	4.4	6	3.6	4.5	20.8	21.1
2020/5/26	4.4	5.3	3.6	4.6	20.9	21.3
2020/5/27	4.4	5.3	3.8	4.4	21.1	21.5
2020/5/28	4.7	5.6	3.9	4.5	21.3	21.5
2020/5/29	4.3	5.3	3.6	4.6	21.2	21.5
2020/5/30	4.3	5.1	3.6	4.4	21.2	21.4
2020/5/31	4.3	4.9	3.6	4.4	21.1	21.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥  
師(三)級藥師 10906010905  
藥學部 溫雅如  
師(三)級藥師 1090601430