

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109年04月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/4/1	3.9	4.4	3.9	4.8	4.4	5.1	-19.9	-19.7	-81.5	-80.3	19.2	20.1
2020/4/2	4.1	4.4	4.2	4.8	4.6	5.0	-19.9	-19.7	-81.6	-80.3	19.3	19.4
2020/4/3	4.1	4.4	4.2	4.8	4.6	4.9	-19.9	-19.7	-81.5	-80.1	19.2	19.3
2020/4/4	4.1	4.4	4.2	4.8	4.6	4.9	-19.9	-19.7	-81.5	-80.3	18.8	19.2
2020/4/5	4.1	4.4	4.2	4.8	4.6	4.9	-19.9	-19.8	-81.7	-80.3	18.7	18.8
2020/4/6	3.9	4.5	3.9	4.8	4.5	5.1	-20.7	-19.6	-81.7	-80.3	18.6	19.8
2020/4/7	3.9	4.4	4.0	4.8	4.4	5.2	-20.7	-20.3	-81.6	-80.3	18.9	19.7
2020/4/8	4.0	4.6	4.0	4.8	4.3	5.0	-20.9	-20.4	-81.7	-80.3	19.0	19.9
2020/4/9	3.9	4.7	4.0	4.9	4.3	4.9	-20.9	-20.3	-81.5	-80.3	19.1	19.8
2020/4/10	3.9	4.5	4.0	4.9	4.2	4.8	-20.9	-20.4	-81.5	-80.3	19.0	19.8
2020/4/11	4.2	4.5	4.2	4.8	4.3	4.8	-20.9	-20.4	-81.5	-80.2	19.1	19.3
2020/4/12	4.2	4.5	4.2	4.8	4.3	4.8	-20.9	-20.6	-81.5	-80.3	18.6	19.1
2020/4/13	3.9	4.6	4.0	4.8	4.2	4.8	-20.9	-20.4	-81.6	-80.1	18.4	19.2
2020/4/14	3.9	4.5	3.9	4.8	4.2	4.8	-20.9	-20.5	-81.5	-80.3	18.6	20.2
2020/4/15	3.9	4.6	3.9	4.8	4.3	4.9	-20.9	-20.4	-81.7	-80.3	19.8	20.8
2020/4/16	4.0	4.7	3.9	5.0	4.3	5.0	-20.9	-20.4	-81.5	-80.1	19.9	21.2
2020/4/17	4.0	4.9	3.9	4.8	4.4	5.1	-20.8	-19.9	-81.7	-80.1	20.3	21.4
2020/4/18	4.6	4.9	4.3	4.8	4.4	4.9	-20.7	-20.3	-81.7	-80.3	19.9	20.4
2020/4/19	4.6	4.9	4.3	4.8	4.4	4.9	-20.8	-20.4	-81.7	-80.3	19.8	19.9
2020/4/20	3.9	4.9	4.0	4.8	4.4	5.1	-20.8	-20.0	-81.7	-80.1	19.4	20.2
2020/4/21	3.9	4.9	4.0	4.8	4.4	5.0	-21.2	-20.4	-81.7	-80.3	19.2	20.2
2020/4/22	4.0	5.6	3.9	4.8	4.4	5.7	-21.2	-20.7	-81.7	-80.3	19.3	20.4
2020/4/23	4.2	5.1	3.9	4.7	4.5	5.2	-20.9	-20.4	-81.7	-80.1	19.3	20.1
2020/4/24	4.3	5.3	4.0	5.0	4.4	5.2	-20.9	-20.3	-81.7	-80.3	19.2	20.2
2020/4/25	4.8	5.3	4.3	4.8	4.4	4.9	-20.9	-20.4	-81.7	-80.3	19.1	19.2
2020/4/26	4.8	5.2	4.3	4.8	4.4	4.9	-20.9	-20.4	-81.7	-80.3	19.1	19.4
2020/4/27	4.4	5.4	4.0	4.9	4.4	5.3	-20.8	-20.4	-81.7	-80.3	19.4	20.3
2020/4/28	4.3	5.3	4.1	4.8	4.4	5.3	-20.7	-20.2	-81.6	-78.9	19.2	19.8
2020/4/29	4.4	5.3	4.1	4.8	4.4	5.2	-20.7	-20.0	-81.0	-80.1	18.9	19.7
2020/4/30	4.5	5.3	3.9	5.1	4.4	5.2	-20.9	-20.0	-81.1	-80.2	18.9	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 溫雅如  
師(三)級藥師 10905010920  
藥學部 莊承芳  
師(三)級藥師 10905010945