

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109年03月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/3/1	4.4	4.8	4.3	4.9	4.8	5.3	-20.3	-20.1	-80.4	-79.4	18.6	18.7
2020/3/2	4.2	4.7	4.1	4.8	4.5	5.2	-20.4	-19.2	-80.4	-79.3	18.4	19.2
2020/3/3	4.2	4.7	4.2	4.9	4.4	5.5	-20.4	-19.6	-80.4	-79.3	18.6	19.4
2020/3/4	2.4	4.7	4.2	4.9	4.3	5.1	-20.2	-19.6	-80.4	-79.2	18.8	19.5
2020/3/5	4.2	4.7	4.1	5.1	4.4	5.1	-21.2	-19.9	-80.4	-79.2	18.7	19.4
2020/3/6	4.3	4.7	4	4.8	4.4	5.1	-21.4	-20.4	-80.4	-79.3	18.7	19.5
2020/3/7	4.3	4.8	4.3	4.8	4.5	5.1	-21.7	-20.9	-80.4	-79.4	18.6	18.8
2020/3/8	4.3	4.7	4.3	4.8	4.6	5.1	-21.5	-20.9	-80.4	-79.3	18.6	19
2020/3/9	4.3	4.9	4.1	4.9	4.3	5.2	-21.5	-20.9	-80.6	-79.4	18.9	19.7
2020/3/10	4.3	4.8	4.2	4.9	4.4	5.2	-21.7	-19.4	-80.4	-79.3	18.2	19.3
2020/3/11	4.3	4.8	4.1	4.9	4.4	5.1	-20.9	-20.2	-80.6	-79.2	17.7	19.5
2020/3/12	4.3	5	4.1	4.8	4.4	5.2	-20.4	-20	-80.6	-79.3	19.2	20.3
2020/3/13	4.2	4.8	4.1	4.8	4.4	5.1	-20.4	-19.9	-80.4	-79.1	19.1	20.4
2020/3/14	4.3	4.7	4.3	4.8	4.4	4.9	-20.4	-19.9	-80.6	-79.3	18.3	19
2020/3/15	4.3	4.7	4.3	4.8	4.4	4.9	-20.4	-19.9	-80.6	-79.3	17.9	18.3
2020/3/16	4.3	4.9	4.2	4.9	4.4	5.1	-21	-19.7	-80.5	-79.3	17.9	19.3
2020/3/17	4.2	4.9	4.2	4.9	4.5	5.3	-21.2	-20.7	-80.6	-79.3	19	20.2
2020/3/18	4.2	4.8	4	4.8	4.6	5.3	-21.1	-20.4	-80.6	-79.4	18.6	20.3
2020/3/19	4.2	4.8	4.1	4.8	4.7	5.3	-21.1	-20.5	-80.6	-79.1	18.8	20.4
2020/3/20	4.2	4.8	4	4.9	4.6	5.4	-20.9	-20.1	-80.6	-79.3	19.7	20.5
2020/3/21	4.3	4.8	4.2	4.8	4.6	5.1	-21	-20.6	-80.6	-79.1	19	20.1
2020/3/22	4.3	4.7	4.2	4.8	4.5	5	-21.1	-20.7	-80.6	-79.4	18.7	18.9
2020/3/23	4.1	4.7	4	4.8	4.5	5.2	-21.1	-20.5	-80.6	-79.4	18.6	19.6
2020/3/24	4.2	4.7	3.9	4.8	4.6	5.3	-21.2	-20.6	-80.6	-79.2	18.6	19.5
2020/3/25	4.2	4.7	3.9	4.8	4.4	5.1	-21.2	-20.7	-80.6	-79.4	18.6	19.7
2020/3/26	4.2	4.7	3.9	4.8	4.3	5.1	-21.2	-20.4	-80.6	-79.3	18.7	19.7
2020/3/27	4.2	4.7	3.9	5.1	4.3	5.4	-21.2	-20.7	-80.6	-79.1	18.6	19.6
2020/3/28	4.3	4.7	4.2	4.7	4.5	4.9	-21.2	-20.8	-80.6	-79.3	18.4	18.8
2020/3/29	4.3	4.6	4.2	4.7	4.5	4.9	-21.2	-20.8	-80.4	-79.3	18.4	18.4
2020/3/30	4.2	4.7	3.9	4.7	4.5	5.4	-21.2	-20.6	-80.6	-79.2	18.4	19.8
2020/3/31	4.1	4.7	3.9	4.7	4.3	5.3	-20.9	-20.2	-80.4	-79.4	19.1	20.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:
 溫度均在正常範圍內
 藥學部 邱保祥 109
 師(三) 藥師 1345
 藥學部 邱保祥 109