

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109年02月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/2/1	4.5	4.9	4.4	5.1	4.7	5.2	-20.7	-20.4	-81.6	-79.7	19	19.6
2020/2/2	4.5	4.9	4.4	5.1	4.7	5.2	-20.7	-20.4	-81.8	-80	19.6	19.9
2020/2/3	4.5	5.1	4.3	5.1	4.4	5.3	-20.6	-18.8	-81.6	-80.2	20.2	20.9
2020/2/4	4.2	5.1	4.2	5.1	4.7	5.3	-20.7	-19.7	-81.7	-79.9	20.1	21.4
2020/2/5	4.1	4.8	4.3	4.9	4.7	5.2	-20.8	-20.4	-81.5	-79.7	19.7	20.3
2020/2/6	4.1	4.8	4.2	5	4.5	5.2	-20.9	-20	-81.6	-80	19.8	21.5
2020/2/7	4.1	5.1	4.3	5.6	4.5	5.2	-20.9	-20.7	-81.5	-79.9	19.5	20.3
2020/2/8	4.3	4.7	4.6	5.1	4.7	5.2	-20.9	-20.6	-81.6	-79.6	19.2	19.8
2020/2/9	4.3	4.7	4.6	5.1	4.7	5.2	-20.9	-20.6	-81.6	-79.3	19.8	19.9
2020/2/10	4.2	4.7	4.4	5.1	4.5	5.2	-20.9	-20.2	-81.7	-79.4	19.8	20.9
2020/2/11	4.2	4.9	4.3	5.1	4.5	5.3	-20.6	-18.7	-81.5	-79	20.3	22.1
2020/2/12	4.3	5.2	4.4	5.1	4.7	5.3	-20.4	-20	-81.5	-79	20.5	21.4
2020/2/13	4.4	5.3	4.3	5.3	4.7	5.3	-20.4	-19.8	-81.5	-79.1	20.1	21.3
2020/2/14	4.3	5.3	4.3	5.3	4.8	5.3	-20.7	-20	-81.4	-78.9	19.9	20.8
2020/2/15	4.7	5.3	4.7	5.3	4.8	5.3	-20.7	-20.4	-81.4	-79.3	19.8	20.2
2020/2/16	4.8	5.3	4.6	5.2	4.8	5.3	-20.8	-20.4	-81.4	-79.6	19.2	19.8
2020/2/17	4.3	5.2	4.3	5.2	4.6	5.3	-20.9	-20.2	-81.4	-79	18.9	20.1
2020/2/18	4.4	5.2	4.2	5.2	4.6	5.2	-20.9	-20.3	-81.3	-79.3	18.7	19.6
2020/2/19	4.2	5.2	4.3	4.9	4.6	5.3	-20.7	-20.2	-81.5	-79.6	18.8	19.7
2020/2/20	4.3	5.2	4.3	4.9	4.6	5.3	-20.8	-19.9	-81.5	-79.4	19.2	19.9
2020/2/21	4.3	5	4.3	5.1	4.5	5.3	-20.8	-20	-81.5	-79.5	19.3	20.2
2020/2/22	4.4	4.8	4.3	4.9	4.9	5.3	-20.8	-20.5	-81.5	-79.4	19.4	19.5
2020/2/23	4.4	4.8	4.3	4.9	4.8	5.3	-20.9	-20.4	-81.5	-79.2	18.7	19.4
2020/2/24	4.3	4.9	4.2	4.9	4.5	5.3	-20.9	-20.4	-81.5	-79.4	18.3	19.4
2020/2/25	4.3	4.8	4.2	4.9	4.6	5.4	-20.8	-20.2	-81.5	-79.6	18.8	19.4
2020/2/26	4.4	4.9	4.1	4.8	4.7	5.4	-20.7	-19.8	-81.5	-79.7	18.8	19.6
2020/2/27	4.4	5.1	4.1	4.9	4.6	5.4	-20.6	-19.8	-81.5	-79.6	18.8	19.9
2020/2/28	4.4	4.8	4.3	4.8	4.8	5.3	-20.3	-20	-81.5	-79.5	18.8	19
2020/2/29	4.4	4.8	4.3	4.9	4.8	5.3	-20.4	-20	-81.5	-79.3	18.6	18.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 1090302  
藥學部 溫雅如 1090302