

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109年01月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/1/1	3.8	4.2	4.3	4.8	4.7	5.1	-19.9	-19.7	-81.8	-79.7	19.6	19.9
2020/1/2	3.3	4.6	4.2	5.1	4.4	5.3	-20.4	-18.9	-81.8	-79.8	19.5	20.3
2020/1/3	3.4	4.6	4.4	5.1	4.5	5.1	-20.6	-20	-81.8	-79.8	19.6	20.3
2020/1/4	3.7	4.5	4.6	5.1	4.5	4.9	-20.6	-20.2	-81.7	-79.6	19.7	19.9
2020/1/5	3.7	4.4	4.6	5.1	4.5	4.9	-20.6	-20.2	-81.8	-80	19.5	19.7
2020/1/6	3.3	4.6	4.4	5.1	4.4	5	-20.6	-19.9	-81.8	-79.8	19.5	20.3
2020/1/7	3.6	4.6	4.4	5.1	4.5	5.2	-20.3	-19.7	-81.8	-80.1	19.7	20.4
2020/1/8	3.1	4.3	4.4	5.1	4.4	5	-20.4	-19.8	-81.7	-79.8	19.7	20.3
2020/1/9	2.9	4.4	4.3	5.1	4.4	5.1	-20.4	-19.7	-81.8	-80.1	19.5	20.3
2020/1/10	2.8	4.4	4.4	5.3	4.4	4.9	-20.2	-19.6	-81.8	-80	19.7	20.4
2020/1/11	3.1	4.4	4.6	5.1	4.5	4.9	-20.2	-19.9	-81.8	-80.1	19.7	19.9
2020/1/12	3.1	4.3	4.6	5.1	4.5	4.9	-20.3	-19.9	-81.8	-79.8	19.3	19.7
2020/1/13	2.9	4.5	4.4	5.1	4.4	5.1	-20.7	-19.9	-81.8	-79.6	19.3	21.2
2020/1/14	2.8	4.8	4.5	5.2	4.4	6.1	-20.9	-19	-81.7	-80.1	20.3	21.1
2020/1/15	3.7	4.7	4.3	5.2	4.4	5	-20.9	-17.8	-81.8	-80.1	19.6	21
2020/1/16	3.7	4.5	4.2	4.9	4.4	5.1	-19.9	-18.3	-81.8	-80.1	19.2	19.9
2020/1/17	3.9	4.4	4.2	6	4.4	5.1	-19.6	-18.6	-81.7	-79.6	19	19.6
2020/1/18	4.2	4.4	4.4	5.1	4.5	5.1	-19.6	-19.3	-81.8	-79.8	18.7	19
2020/1/19	4.2	4.4	4.4	4.9	4.5	5.1	-19.5	-19.3	-81.7	-79.5	18.6	19
2020/1/20	3.9	4.7	4.2	4.9	4.4	5.2	-19.7	-19	-81.8	-80	18.9	19.8
2020/1/21	4.2	4.9	4.2	4.9	4.4	5	-20.2	-18.9	-81.8	-80.1	19.1	20.1
2020/1/22	4.2	5.1	4.2	4.9	4.4	5	-20.4	-18.7	-81.7	-79.5	19.4	20.3
2020/1/23	4.3	4.7	4.4	4.9	4.5	5	-20.3	-20.1	-81.8	-79.8	19.4	19.8
2020/1/24	4.3	4.7	4.4	4.9	4.5	5	-20.3	-20	-81.7	-79.9	19.5	19.9
2020/1/25	4.3	4.7	4.4	4.9	4.5	5	-20.3	-20	-81.7	-79.4	19.9	20.1
2020/1/26	4.3	4.7	4.4	4.9	4.5	5	-20.3	-20	-81.7	-79.6	19.9	20.1
2020/1/27	4.3	4.7	4.4	4.9	4.5	5	-20.3	-20	-81.7	-79.9	19.7	19.9
2020/1/28	4.3	4.7	4.4	5	4.5	5	-20.3	-20	-81.7	-80	19.5	19.9
2020/1/29	4.3	4.7	4.3	5.1	4.3	5	-20.4	-20	-81.8	-79.5	19.5	20.3
2020/1/30	4.3	4.8	4.3	5	4.3	5.1	-20.9	-20.3	-81.7	-79.6	19	20.2
2020/1/31	4.3	5.1	4.3	5.1	4.5	5.3	-20.9	-20.4	-81.7	-79.9	18.6	20

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 溫度均在正常範圍內 109
203
1015
109
205
108
藥學部 邱保祥
師(三)級藥師
藥學部 溫雅如
師(二)級藥師