

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年12月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/12/1	4	4.4	4.6	4.9	4.9	5.3	-19.9	-19.6	-80.6	-79.5	19.8	20.1
2019/12/2	3.5	4.5	4.3	5.1	4.4	5.3	-20.7	-18.8	-80.6	-79.6	19.7	20.3
2019/12/3	3.4	4.4	4.3	5.1	4.4	5.4	-20.6	-19	-80.6	-79.5	19.9	20.7
2019/12/4	3.4	4.9	4.3	5.1	5	5.4	-20.4	-19.4	-80.6	-79.6	20.5	20.7
2019/12/5	3.4	4.6	4.4	5.1	4.4	5.3	-20.4	-19.1	-80.6	-79.6	20.4	20.7
2019/12/6	3.7	4.6	4.4	5.1	4.4	5.3	-20.7	-19.6	-80.6	-79.6	20.5	20.8
2019/12/7	4.1	4.5	4.6	4.9	4.7	5.2	-19.9	-19.7	-80.6	-79.4	19.9	20.5
2019/12/8	4.1	4.7	4.6	5.1	4.8	5.3	-19.9	-19.7	-80.6	-79.6	19.7	19.9
2019/12/9	3.8	5.1	4.3	5	4.4	5.3	-19.9	-19.4	-80.6	-79.5	19.6	19.9
2019/12/10	3.9	4.6	4.3	5.1	4.3	5.2	-19.6	-19.1	-80.6	-79.6	19.9	20.2
2019/12/11	3.7	4.5	4.4	5.1	4.4	5.2	-20.6	-19.3	-80.6	-79.3	19.9	20.4
2019/12/12	3.7	4.6	4.4	5.1	4.4	5.2	-20.6	-20.1	-80.6	-79.4	20.1	20.4
2019/12/13	3.7	4.5	4.3	5.1	4.4	5.3	-20.4	-19.8	-80.6	-79.4	19.9	20.3
2019/12/14	3.9	4.5	4.7	5.2	4.8	5.3	-20.2	-19.9	-80.7	-79.6	19.9	20.1
2019/12/15	4	4.4	4.7	5.1	4.8	5.2	-20.2	-19.9	-80.7	-79.3	19.7	19.9
2019/12/16	3.7	4.5	4.4	5.1	4.4	5.2	-20.6	-19.9	-80.6	-79.3	19.7	20.1
2019/12/17	3.7	4.5	4.3	6.3	4.3	5.1	-20.7	-18.4	-80.9	-79.6	19.8	20.3
2019/12/18	3.5	4.7	4.6	5.4	4.3	5.1	-19.7	-19.6	-80.9	-79.6	19.9	20.3
2019/12/19	3.4	4.4	4.4	5.2	4.3	5	-19.7	-19.5	-80.9	-79.7	19.7	20.1
2019/12/20	3.7	4.5	4.5	5.2	4.3	4.8	-19.7	-19.4	-80.9	-79.3	19.7	20.1
2019/12/21	4	4.5	4.7	5.1	4.3	4.7	-19.6	-19.6	-80.9	-79.4	19.6	19.9
2019/12/22	3.9	4.5	4.6	5.1	4.3	4.8	-19.7	-19.6	-80.9	-79.6	19.3	19.6
2019/12/23	3.7	4.7	4.4	5.1	4.2	4.8	-19.8	-19.5	-80.8	-79.7	19.3	19.7
2019/12/24	3.7	4.6	4.4	5.1	4.3	4.9	-19.8	-19.3	-80.8	-79.8	19.4	19.9
2019/12/25	3.1	4.4	4.3	5.1	4.2	4.9	-19.6	-19.1	-80.8	-79.4	19.6	19.8
2019/12/26	3.3	4.4	4.4	5.1	4.1	4.9	-20.1	-19.1	-80.9	-79.4	19.6	20.1
2019/12/27	3.4	4.6	4.2	4.9	4	4.9	-20.4	-19.2	-80.7	-79.6	19.4	19.7
2019/12/28	3.8	4.4	4.3	4.8	4.4	4.9	-20.4	-19.9	-80.6	-79.1	19.4	20.3
2019/12/29	3.8	4.4	4.3	4.8	4.5	4.9	-20.3	-20	-80.6	-79.3	20.3	20.8
2019/12/30	3.7	4.4	4.2	4.8	4.1	5.1	-20.3	-19.6	-80.7	-79.3	20.8	21.6
2019/12/31	3.6	4.5	4.2	4.8	4.3	5.1	-20.3	-19.6	-80.7	-79.3	20.4	20.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部
師(三)級藥師 邱保祥
藥學部
師(三)級藥師 溫雅如