

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年11月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/11/1	4.3	5	3.9	4.9	4.2	4.8	-19.7	-18	-81.4	-79.8	19.7	20.2
2019/11/2	4.3	4.8	4.6	4.9	4.3	5	-19.7	-19.6	-81.3	-79.7	19.5	19.9
2019/11/3	4.3	4.7	4.6	4.8	4.3	4.9	-19.7	-19.5	-81.3	-79.6	19.4	19.5
2019/11/4	4.3	5.1	4.3	4.9	4.2	4.9	-19.7	-19.1	-81.4	-79.8	19.3	20.1
2019/11/5	4.3	5.1	4.3	5	4.2	4.9	-19.7	-19.3	-81.2	-79.7	19.4	20.2
2019/11/6	4.3	5.1	4.3	5.1	4.2	4.9	-19.7	-18.8	-81.4	-79.6	19.4	20.1
2019/11/7	4.4	5.1	4.3	5.1	4.2	4.9	-19.6	-19	-81.4	-79.7	19.4	20.2
2019/11/8	4.3	5.4	4.3	5.1	4.3	4.9	-19.7	-19.1	-81.2	-79.4	19.3	19.9
2019/11/9	4.3	4.8	4.5	5.1	4.3	4.8	-19.7	-19.5	-81.4	-79.8	19.2	19.7
2019/11/10	4.3	4.8	4.5	5.1	4.4	4.8	-19.7	-19.6	-81.5	-79.7	19.1	19.8
2019/11/11	4.3	5.3	4.3	5.1	4.2	4.9	-20.6	-19.6	-81.5	-79.9	18.9	19.7
2019/11/12	4	5.3	4.3	4.9	4.2	4.9	-20.6	-19.8	-81.5	-79.9	19.2	20.1
2019/11/13	4	4.5	4.4	5.1	4.2	5.1	-20.2	-19.8	-81.5	-79.6	19.4	20.3
2019/11/14	3.9	4.5	4.3	5.1	4.4	5	-20.2	-19.6	-81.5	-79.9	19.5	20.3
2019/11/15	3.9	4.7	4.4	5.1	4.4	5	-19.8	-19.4	-81.5	-79.7	19.5	20.3
2019/11/16	4.2	4.7	4.7	5.1	4.5	5	-19.6	-19.4	-81.5	-79.6	19.6	20.2
2019/11/17	4.2	4.7	4.7	5.1	4.5	5	-19.6	-19.4	-81.5	-79.9	19.5	19.6
2019/11/18	4.2	4.8	4.4	5.3	4.4	5.1	-19.3	-19	-81.5	-79.6	19.4	20.3
2019/11/19	4.2	4.9	4.4	5.1	4.4	5.1	-19.3	-18.5	-81.5	-79.8	19.5	20.3
2019/11/20	4.2	5.1	4.4	5.1	4.4	5.2	-19.1	-17.9	-81.5	-79.8	19.6	20.4
2019/11/21	4.4	5	4.4	5.1	4.5	5.2	-19.7	-18.8	-81.4	-79.6	19.6	21.3
2019/11/22	4.2	4.9	4.4	5.1	4.5	5.2	-19.7	-18.8	-81.5	-79.6	19.7	20.5
2019/11/23	4.2	4.8	4.7	5.2	4.7	5.4	-18.9	-18.7	-81.5	-79.7	19.7	20.2
2019/11/24	4.2	4.7	4.7	5.1	4.6	5.1	-19.1	-18.8	-81.5	-79.8	19.5	19.7
2019/11/25	4.2	4.9	4.3	5.1	4.5	5.3	-19.1	-18.7	-81.5	-79.9	19.4	20.2
2019/11/26	4.2	4.9	4.3	4.9	4.5	5.1	-19.1	-18.7	-81.5	-79.8	19.5	20.3
2019/11/27	4.3	4.9	4.3	5.1	4.5	5.1	-19.1	-18.7	-81.5	-79.6	19.6	20.3
2019/11/28	4.2	4.8	4.4	5	4.5	5.2	-19.1	-18.8	-81.5	-79.9	19.5	20.2
2019/11/29	2.8	4.9	4.4	5.1	4.7	5.1	-19.8	-19	-81.5	-79.9	19.4	20.1
2019/11/30	3.3	4.2	4.6	5.1	4.7	5.2	-19.9	-19.6	-81.5	-79.9	19.5	20.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥 108  
師(三級藥師) 邱保祥 1202  
藥學部 溫推如 108  
師(三級藥師) 溫推如 1202