

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年10月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2019/10/1	4.4	5	4.8	5.9	20.6	21.3
2019/10/2	4.2	4.9	5.5	6.4	21.1	22.2
2019/10/3	4.3	5.6	5.7	6.2	21.9	22.6
2019/10/4	4.7	5.6	4.9	5.9	21.9	22.6
2019/10/5	4.7	5.4	5.3	5.9	21.9	22.3
2019/10/6	4.7	5.4	5.5	5.8	21.6	22.1
2019/10/7	4.7	5.4	5.5	5.9	21.3	21.8
2019/10/8	4.7	5.4	5.4	5.9	21.5	22.1
2019/10/9	4.7	5.4	5.6	6.1	21.7	22.3
2019/10/10	4.8	5.4	5.7	5.9	21.7	22.1
2019/10/11	4.8	5.4	5.6	5.9	21.3	21.7
2019/10/12	4.8	5.4	5.6	5.9	20.9	21.3
2019/10/13	4.8	5.4	5.6	5.9	20.8	21.1
2019/10/14	4.8	6.2	5.4	5.8	20.6	21.3
2019/10/15	5.6	6.2	5.3	5.8	21.1	21.7
2019/10/16	5.3	6.2	5.5	6.1	21.3	22
2019/10/17	5.3	5.9	5.7	6.1	21.5	21.9
2019/10/18	5.3	6	5.7	6.2	21.7	22.6
2019/10/19	5.5	5.9	5.8	6.1	21.9	22.4
2019/10/20	5.5	5.9	5.7	6	21.3	21.9
2019/10/21	5.4	6.1	5.7	6.1	21.1	21.8
2019/10/22	5.4	6.2	5.7	6.1	21.3	22
2019/10/23	5.7	6.3	5.7	6.2	21.7	22.3
2019/10/24	5.2	6.1	5.7	6.6	21.7	22.3
2019/10/25	5.1	5.8	5.6	6	21.7	22.2
2019/10/26	5.1	5.7	5.6	5.9	21.7	22
2019/10/27	5.1	5.6	5.6	5.9	21.2	21.6
2019/10/28	4.9	5.8	5.6	6.1	21.1	21.3
2019/10/29	4.9	5.7	5.5	6.2	21.2	21.6
2019/10/30	4.9	5.6	5.4	5.9	21.3	21.7
2019/10/31	4.9	5.7	5.3	6.1	21.4	21.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 溫度均在正常範圍內

藥學部 邱保祥 (08101)

藥學部 溫雅如 (108110)