

臺北榮民總醫院藥學部
臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 108年09月

	六號冰箱		室溫	
	序號	1436-0086	備用	1118-0011
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2019/9/1	3.6	3.9	19.6	19.8
2019/9/2	3.6	4.7	19.8	20.1
2019/9/3	3.5	4	19.9	20.5
2019/9/4	3.6	4	20.1	20.4
2019/9/5	3.7	4.2	20.3	20.7
2019/9/6	3.7	5.2	20.4	20.8
2019/9/7	3.9	4.2	20.1	20.7
2019/9/8	3.9	4.2	19.9	20.1
2019/9/9	3.9	4.2	19.9	20.2
2019/9/10	3.7	4.2	19.9	20.2
2019/9/11	3.6	4.2	19.7	20.3
2019/9/12	3.4	4.2	19.9	20.7
2019/9/13	3.9	4.2	19.9	20.3
2019/9/14	3.9	4.2	19.7	19.9
2019/9/15	3.9	4.2	19.6	19.9
2019/9/16	3.7	4.3	19.5	20.1
2019/9/17	3.6	4.2	19.9	20.7
2019/9/18	3.9	4.2	19.7	20.6
2019/9/19	3.7	4.2	20.2	20.8
2019/9/20	3.9	4.2	20.1	21
2019/9/21	3.9	4.2	20.4	20.9
2019/9/22	3.9	4.2	20.3	20.8
2019/9/23	3.9	4.2	19.9	20.7
2019/9/24	3.9	4.2	20.2	20.7
2019/9/25	3.9	4.1	19.9	20.7
2019/9/26	3.9	4.2	19.9	20.8
2019/9/27	4	4.3	20.1	20.3
2019/9/28	4.1	4.3	19.7	20.2
2019/9/29	4.1	4.3	19.5	19.7
2019/9/30	4.1	4.3	19.5	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師 : **藥學部邱保祥** 108
(三)級藥師 100
藥學部溫雅如 108
(三)級藥師 100