

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年08月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/8/1	4.2	4.7	4.4	5	4.3	4.9	-18.5	-17.8	-81.3	-79.7	19.2	19.8
2019/8/2	4	4.6	4.5	4.9	4.2	4.9	-18.7	-17.9	-81.4	-79.7	18.9	19.8
2019/8/3	4.2	4.6	4.5	4.9	4.5	4.9	-18.7	-18.7	-81.4	-79.6	18.8	19.1
2019/8/4	4.1	4.6	4.5	4.9	4.5	4.9	-18.7	-18.7	-81.4	-79.3	18.8	18.9
2019/8/5	3.9	4.8	4.5	5.1	4.2	4.9	-18.7	-16.1	-81.3	-79.3	18.6	20
2019/8/6	3.9	4.7	4.6	5.1	4.3	4.9	-18.8	-17.8	-81.4	-79.7	18.9	20.3
2019/8/7	3.8	4.8	4.5	5.1	4.2	4.9	-18.8	-18.2	-81.4	-79.4	18.6	20
2019/8/8	4.2	4.8	4.5	5.1	4.2	4.9	-18.9	-18.1	-81.4	-79.7	18.6	19.6
2019/8/9	4.3	4.8	4.5	4.9	4.4	4.9	-18.3	-18.2	-81.4	-79.7	18.3	18.7
2019/8/10	4.3	4.8	4.4	4.9	4.4	4.9	-18.3	-18.3	-81.4	-79.7	18.1	18.3
2019/8/11	4.3	4.8	4.5	4.9	4.4	4.9	-18.3	-18.3	-81.4	-79.6	18	18.2
2019/8/12	4	4.7	4.4	4.9	4.2	4.9	-18.3	-18.1	-81.4	-79.7	17.9	19.9
2019/8/13	4	4.8	4.4	5	4.3	4.9	-18.3	-17.7	-81.4	-79.2	19.2	20.4
2019/8/14	3.9	5	4.5	5.1	4.2	5	-18.1	-17.6	-81.4	-79.7	19.4	20.6
2019/8/15	3.8	4.7	4.4	5	4.3	4.9	-18.1	-17.7	-81.2	-79.4	18.7	20.1
2019/8/16	3.8	4.6	4.4	4.9	4.2	4.9	-18.1	-17.8	-81.4	-79.5	18.4	19.8
2019/8/17	4.2	4.6	4.4	4.9	4.5	4.9	-18.1	-18.1	-81.4	-79.6	18.8	19.1
2019/8/18	4.2	4.6	4.4	4.8	4.4	4.9	-18.3	-18.1	-81.4	-79.7	18.2	18.8
2019/8/19	3.9	4.6	4.3	4.9	4.4	5.3	-18.7	-18.1	-81.4	-79.3	18.4	19.9
2019/8/20	3.9	4.7	4.3	5.3	4.6	5.2	-18.9	-18.7	-81.4	-79.6	18.6	19.8
2019/8/21	4.1	4.8	4.4	5.1	4.5	5.2	-18.9	-18.1	-81.3	-79.2	18.8	19.8
2019/8/22	4.1	4.8	4.4	5.1	4.4	5.4	-18.3	-17.9	-81.4	-79.7	18.7	19.8
2019/8/23	4	4.7	4.4	5.1	4.6	5.4	-19.8	-18.1	-81.3	-79.8	18.3	19.8
2019/8/24	4.3	4.7	4.4	4.9	4.6	5.1	-19.9	-19.7	-81.4	-79.2	18.3	18.8
2019/8/25	4.3	4.7	4.4	4.9	4.6	5	-19.9	-19.7	-81.3	-79.7	18.7	18.8
2019/8/26	4	4.7	4.4	5.1	4.6	5.2	-19.9	-19.6	-81.4	-79.7	18.6	19.8
2019/8/27	4.1	4.8	4.4	5.1	4.6	5.3	-19.9	-18.4	-81.2	-79.3	18.6	19.8
2019/8/28	4	4.7	4.3	5.1	4.5	5.2	-19.3	-18.7	-81.3	-79.6	18.8	19.8
2019/8/29	4.2	4.9	4.3	5.2	4.4	5.1	-19.6	-18.8	-81.4	-79.5	18.8	19.9
2019/8/30	4.1	4.7	4.3	4.9	4.4	5.1	-19.6	-18.8	-81.2	-79.5	18.8	20.1
2019/8/31	4.2	4.7	4.3	4.8	4.5	5.1	-19.1	-18.9	-81.4	-79.7	18.4	18.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保木 1080902
 師(三級藥師)
 藥學部 溫雅如 1080902
 師(二級藥師)