

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年08月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2019/8/1	4.7	5.8	4.8	5.2	21.3	21.7
2019/8/2	4.7	5.6	4.8	5.3	21.2	21.7
2019/8/3	4.7	5.3	4.8	5.3	21.2	21.6
2019/8/4	4.7	5.3	4.9	5.2	20.8	21.2
2019/8/5	4.6	5.3	4.8	5.4	20.6	21.3
2019/8/6	4.6	5.3	4.9	5.4	21.1	21.6
2019/8/7	4.6	5.6	4.9	5.4	21.1	22
2019/8/8	4.7	5.4	4.8	5.6	21.8	22.1
2019/8/9	4.7	5.3	5.1	5.3	21	21.8
2019/8/10	4.7	5.3	4.9	5.2	20.6	21
2019/8/11	4.7	5.3	4.9	5.3	20.3	20.6
2019/8/12	4.7	5.5	4.9	5.4	20.1	21.2
2019/8/13	4.8	5.4	4.9	5.6	20.9	21.8
2019/8/14	4.8	5.5	5.1	5.6	21.4	21.9
2019/8/15	4.9	5.5	4.9	5.4	21.3	21.9
2019/8/16	4.8	5.6	4.4	6.2	21.1	21.8
2019/8/17	4.8	5.4	4.9	5.2	21.1	21.6
2019/8/18	4.8	5.4	4.9	5.2	20.5	21.1
2019/8/19	4.8	5.4	4.8	5.4	20.3	20.9
2019/8/20	4.8	5.5	4.8	5.2	20.6	21.1
2019/8/21	4.8	6.3	4.8	6.1	20.9	21.6
2019/8/22	4.8	6.1	4.9	5.4	20.9	21.3
2019/8/23	4.9	5.6	4.9	5.5	20.7	21.3
2019/8/24	4.9	5.5	4.9	5.2	20.7	21.1
2019/8/25	4.9	5.5	4.9	5.2	20.4	20.8
2019/8/26	4.9	5.6	4.9	5.8	20.3	20.9
2019/8/27	4.8	5.6	5.5	5.9	20.6	21.1
2019/8/28	4.7	5.4	5.6	6	20.8	21.3
2019/8/29	4.7	5.6	5.3	5.9	21.1	21.6
2019/8/30	4.4	5.3	5.4	6	21.2	21.7
2019/8/31	4.4	4.9	5.6	5.9	21	21.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 108
師(三)級藥師 0902 藥學部 溫雅如 108
師(三)級藥師 0902