

臺北榮民總醫院藥學部
臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年06月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2019/6/1	4.9	5.4	4.7	4.9	20.8	21.6
2019/6/2	4.9	5.4	4.6	4.9	20.3	20.8
2019/6/3	4.7	5.4	4.5	4.9	20.2	20.9
2019/6/4	4.8	5.6	4.6	5.8	20.7	21.3
2019/6/5	4.9	5.5	5.3	5.8	20.9	21.4
2019/6/6	4.9	5.6	4.6	5.7	21.1	21.7
2019/6/7	5.1	5.6	4.6	5	20.9	21.6
2019/6/8	5.1	5.6	4.6	5	20.4	20.9
2019/6/9	5.1	5.6	4.6	5	20.2	20.4
2019/6/10	5.1	5.6	4.5	5.9	20.1	20.7
2019/6/11	5.1	5.7	4.7	5.3	20.5	21.1
2019/6/12	5.1	5.7	4.7	5.2	20.7	21.4
2019/6/13	5.1	5.7	4.6	7.3	20.7	21.3
2019/6/14	5.1	5.9	4.6	5.6	20.7	21.3
2019/6/15	5.3	5.7	5.3	5.6	20.9	21.1
2019/6/16	5.3	5.6	5.3	5.6	20.7	20.9
2019/6/17	5.3	5.8	5.3	5.8	20.6	21.3
2019/6/18	5.2	5.9	5.2	6.7	21.1	21.7
2019/6/19	5.2	5.8	5.4	5.8	21.3	21.9
2019/6/20	5.2	5.7	5.3	6.1	21.3	21.9
2019/6/21	5.2	5.8	5.3	5.9	21.3	21.7
2019/6/22	5.2	5.7	5.2	5.6	21.1	21.5
2019/6/23	5.2	5.6	5.2	5.5	20.6	21.1
2019/6/24	5.2	5.7	5.2	5.6	20.3	20.9
2019/6/25	5.2	5.7	5.2	5.9	20.6	21.1
2019/6/26	5.1	5.7	5.3	6	20.7	21.2
2019/6/27	5.1	5.7	5.3	5.9	20.8	21.3
2019/6/28	5.1	5.7	5.4	5.7	20.9	21.7
2019/6/29	5.1	5.6	5.4	5.8	20.9	21.3
2019/6/30	5.1	5.6	5.4	5.7	20.7	20.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師 : 藥學部邱保祥
 (三級藥師) 108/06/01

藥學部溫雅如
 (三級藥師) 108/06/01