

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年05月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/5/1	3.6	5.1	2.4	6	2.5	5.7	-20.4	-19.2	-79.9	-79.6	19	19.7
2019/5/2	3.5	4.4	5.4	6.1	4.8	5.5	-20.3	-19.8	-79.9	-79.6	18.7	19.8
2019/5/3	3.4	4.4	5.4	6.2	4.8	5.6	-20.3	-19.7	-79.9	-79.5	18.8	21.1
2019/5/4	3.7	4.3	5.7	6.1	5.1	5.5	-20.1	-19.8	-79.9	-79.4	19.9	20.1
2019/5/5	3.7	4.3	5.7	6.1	5.1	5.4	-20.1	-19.8	-79.9	-79.5	19.8	19.9
2019/5/6	3.3	4.4	5.7	6.2	4.8	5.6	-20.3	-19.8	-79.9	-79.4	19.7	20.9
2019/5/7	3.3	4.4	5.7	6.3	4.7	5.3	-20.4	-19.9	-79.9	-79.5	19.9	21.1
2019/5/8	3.4	4.3	4.7	6.2	4.7	5.4	-20.9	-19.7	-79.9	-79.5	20.1	21.1
2019/5/9	3.3	4.4	4.7	5.4	4.6	5.3	-20.9	-20.4	-79.9	-79.4	20.1	21.2
2019/5/10	3.3	4.4	4.8	5.4	4.5	5.4	-20.9	-20.4	-79.9	-79.4	20.1	21.2
2019/5/11	3.7	4.3	4.8	5.3	4.9	5.3	-20.9	-20.7	-79.9	-79.5	20.1	20.3
2019/5/12	3.7	4.3	4.8	5.2	4.9	5.3	-20.9	-20.7	-79.9	-79.5	19.9	20.1
2019/5/13	3.3	4.3	4.8	5.4	4.5	5.3	-20.9	-20.6	-79.9	-79.4	19.8	20.9
2019/5/14	3.3	4.3	4.8	5.4	4.4	5.3	-21.1	-20.6	-79.9	-79.4	19.9	21.1
2019/5/15	3.4	4.4	4.9	5.5	4.3	5.2	-21.1	-19.8	-79.9	-79.4	20.1	21.2
2019/5/16	3.3	4.5	4.9	5.6	4.4	5.4	-20.8	-19.6	-79.9	-79.4	20.2	21.4
2019/5/17	3.3	4.3	4.9	5.4	4.4	5.1	-20.9	-20.5	-79.9	-79.4	19.3	20.3
2019/5/18	3.7	4.3	4.9	5.3	4.5	5.1	-20.9	-20.8	-79.9	-79.6	18.8	19.3
2019/5/19	3.7	4.3	4.9	5.3	4.5	5.1	-21.1	-20.8	-79.9	-79.6	18.7	18.8
2019/5/20	3.3	4.3	4.6	5.3	4.4	5.1	-21.2	-20.6	-79.9	-79.6	18.6	19.7
2019/5/21	3.3	4.3	4.6	5.1	4.4	5.1	-21.2	-20.6	-79.9	-79.6	18.8	20.1
2019/5/22	3.3	4.3	4.5	5.1	4.4	5.1	-21.2	-20.7	-79.9	-79.5	19.3	20.3
2019/5/23	3.2	4.3	4.3	5.1	4.4	5.2	-21.1	-20.4	-79.9	-79.6	19.1	20.3
2019/5/24	3.2	4.3	4.3	4.8	4.4	5.2	-20.9	-20.4	-79.9	-79.5	19.3	20.4
2019/5/25	3.4	4.1	4.3	4.7	4.5	5.1	-20.9	-20.6	-79.9	-79.6	19.4	19.6
2019/5/26	3.4	4.1	4.3	4.7	4.5	5.1	-20.9	-20.5	-79.9	-79.4	19.4	19.5
2019/5/27	2.9	4.1	4.3	4.8	4.4	5.3	-20.8	-20.3	-79.9	-79.4	19.4	20.6
2019/5/28	3.1	5.9	4.3	4.9	4.4	6.2	-20.9	-19	-79.9	-79.4	19.4	20.6
2019/5/29	3.3	5.1	4.3	4.9	4.4	5.3	-19.7	-18.5	-79.9	-79.5	19.6	20.6
2019/5/30	3.6	4.6	4.3	4.9	4.4	5.3	-19.9	-19.4	-79.9	-79.5	19.3	20.4
2019/5/31	3.1	4.4	4.3	5.1	4.4	5.7	-20.7	-19.6	-79.9	-79.5	19	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內
藥學部溫雅如 080604
藥學部邱保祥 080604