

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年04月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/4/1	3.7	4.6	5.4	6.2	4.3	5.3	-21.4	-20.2	-81.5	-79.6	19.7	20.4
2019/4/2	3.7	4.6	5.5	6.2	4.3	5.3	-21.5	-21.2	-81.5	-77.3	19.6	20.8
2019/4/3	3.7	4.6	5.5	6.2	4.3	5.1	-21.5	-20.9	-81.5	-79.1	19.7	20.9
2019/4/4	4	4.6	5.8	6.2	4.4	4.9	-21.4	-21	-81.5	-79.4	19.7	20
2019/4/5	4	4.6	5.8	6.2	4.4	4.9	-21.4	-21.1	-81.5	-79.5	19.6	19.7
2019/4/6	4	4.6	5.8	6.2	4.4	4.9	-21.4	-21	-81.5	-79.5	19.5	19.6
2019/4/7	4	4.6	5.8	6.2	4.4	4.9	-21.4	-21	-81.5	-79.6	19.5	19.6
2019/4/8	3.7	4.6	5.6	6.3	4.4	5.3	-21.5	-21.1	-81.5	-79.2	19.5	20.8
2019/4/9	3.7	4.7	5.4	6.2	4.3	5.2	-21.5	-20.9	-81.5	-79.1	19.7	20.8
2019/4/10	3.8	4.7	5.4	6.2	4.3	5.1	-21.5	-20.5	-81.5	-79.7	19.4	20.4
2019/4/11	3.7	4.7	5.4	6.2	4.3	5.2	-21.5	-20.8	-81.5	-79.6	19.2	20.6
2019/4/12	3.7	4.7	5.4	6.1	4.3	5.1	-21.5	-20.8	-81.5	-79.7	18.8	20.2
2019/4/13	4.1	4.7	5.7	6.1	4.3	4.7	-21.7	-21.2	-81.5	-79.7	18.9	19.3
2019/4/14	4.1	4.7	5.7	6.1	4.3	4.8	-21.7	-21.2	-81.5	-79.7	18.8	18.9
2019/4/15	3.7	4.7	5.3	6.1	4.3	5.2	-21.7	-20.9	-81.5	-79.7	18.7	20.1
2019/4/16	3.7	4.6	5.3	5.9	4.3	5.1	-21.5	-20.7	-81.5	-79.6	18.9	20.2
2019/4/17	3.7	4.5	5.3	6.1	4.3	5.1	-21.5	-21	-81.5	-79.3	19.1	20.2
2019/4/18	3.6	4.5	5.3	6	4.3	5.1	-21.5	-21.1	-81.5	-79.6	19.2	20.4
2019/4/19	3.6	4.5	5.2	6	4.3	5.4	-21.5	-20.9	-81.5	-79.6	19.2	20.6
2019/4/20	4.1	4.5	5.3	5.9	4.4	4.9	-21.5	-21.2	-81.5	-79.3	19.3	19.5
2019/4/21	4.1	4.5	5.3	5.9	4.4	4.9	-21.5	-21.2	-81.5	-79.7	19.2	19.3
2019/4/22	3.7	4.5	5.3	5.9	4.4	5.4	-21.5	-21	-81.5	-77.8	19.1	20.4
2019/4/23	3.6	4.6	5.3	5.9	4.3	5.4	-21.5	-20.5	-79.9	-79.5	19.2	20.6
2019/4/24	3.8	4.8	5.3	6	4.3	5.2	-20.8	-20.3	-79.9	-79.6	18.8	19.9
2019/4/25	3.8	4.6	5.3	6	3.9	5.4	-20.8	-20.4	-79.9	-79.6	18.6	19.5
2019/4/26	3.9	4.6	4.7	6	3.9	4.6	-20.8	-20.2	-79.9	-79.6	18.3	19.3
2019/4/27	4.1	4.6	5.2	5.8	3.9	4.5	-20.7	-20.4	-79.9	-79.6	18.2	18.5
2019/4/28	4.1	4.6	5.2	5.9	3.9	4.5	-20.7	-20.3	-79.9	-79.6	18	18.6
2019/4/29	3.9	4.6	4.9	5.8	3.7	4.7	-20.7	-19.9	-79.9	-79.6	18.4	19.7
2019/4/30	3.9	4.7	4.8	5.9	3.8	4.5	-20.4	-19.8	-79.9	-79.6	18.7	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 108
藥學部 溫雅如 0501
藥學部 溫雅如