

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年04月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max
2019/4/1	4.3	5.1	4.7	5.2	21.1	21.7
2019/4/2	4.2	5.1	4.7	5.6	21.3	21.9
2019/4/3	4.1	4.9	4.9	5.3	21.6	22.1
2019/4/4	4.1	4.9	4.9	5.3	21.5	21.9
2019/4/5	4.2	4.9	4.9	5.2	21.2	21.5
2019/4/6	4.2	4.9	4.9	5.2	20.9	21.2
2019/4/7	4.2	4.9	4.9	5.2	20.9	21.1
2019/4/8	4.1	4.9	4.9	5.7	20.8	21.8
2019/4/9	4.1	4.9	5	5.6	21.4	22.2
2019/4/10	3.8	5	5.1	5.8	21.5	22.3
2019/4/11	3.8	4.7	4.7	5.5	21.6	22.2
2019/4/12	3.9	4.8	4.7	5.9	21.1	22.1
2019/4/13	3.9	4.7	5.2	5.6	21.1	21.7
2019/4/14	3.9	4.7	5.2	5.5	20.7	21.1
2019/4/15	3.9	4.7	4.4	5.6	20.4	21.3
2019/4/16	3.8	4.9	4.7	5.6	20.8	21.7
2019/4/17	3.9	4.7	4.5	5.4	21.1	21.8
2019/4/18	3.9	4.7	4.5	5.5	21.1	22
2019/4/19	4	4.9	4.5	5.1	21.2	22.2
2019/4/20	4.1	4.9	4.6	5	21.3	21.9
2019/4/21	4.1	4.9	4.6	5	21.1	21.3
2019/4/22	4	4.8	4.6	5.3	20.9	21.8
2019/4/23	4	4.8	4.8	6.6	21.3	22.2
2019/4/24	3.9	4.8	4.8	5.3	21.5	22.2
2019/4/25	3.9	5.3	4.7	5.4	21.1	21.8
2019/4/26	4.3	5.1	4.7	5.2	20.7	21.3
2019/4/27	4.3	5.1	4.7	5.1	20.4	21.1
2019/4/28	4.3	5.1	4.7	5.1	20.1	20.4
2019/4/29	4.3	5.1	4.2	5.6	20.1	21.1
2019/4/30	4.3	5.1	4.2	4.8	20.5	21.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 108 0501
 藥學部 溫雅如
 師(三)級藥師