

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年03月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/3/1	3.9	4.4	4.5	5.6	4.3	4.9	-19.9	-19.7	-81.5	-79.7	19.1	19.2
2019/3/2	3.9	4.4	4.5	5.6	4.3	4.9	-19.9	-19.7	-81.5	-79.7	18.4	19.2
2019/3/3	3.9	4.4	4.6	5.7	4.3	4.9	-19.9	-19.9	-81.5	-79.7	18.1	18.4
2019/3/4	3.6	4.4	4.6	5.7	4.3	5.1	-21.4	-19.8	-81.5	-79.7	18.2	19.3
2019/3/5	3.6	4.5	2.9	5.7	4.3	5.1	-21.4	-20.8	-81.5	-79.7	18	19.4
2019/3/6	3.6	4.5	2.8	5.7	4.3	5.2	-21.2	-16.4	-81.5	-79.2	18.1	19.5
2019/3/7	3.6	4.5	2.8	5.9	4.4	5.2	-21.2	-20.1	-81.5	-79.7	18.1	19.3
2019/3/8	3.6	4.4	5.1	5.8	4.3	5.2	-21.2	-20.5	-81.5	-79.1	18	19
2019/3/9	4	4.4	5.2	5.8	4.3	4.9	-21.2	-20.8	-81.5	-79.7	18	18.2
2019/3/10	4	4.4	5.2	5.8	4.3	4.9	-21.2	-20.7	-81.5	-79.4	18	18.4
2019/3/11	3.6	4.5	5.1	5.8	4.4	5.2	-21.2	-20.7	-81.5	-79.6	18.3	19.7
2019/3/12	3.6	4.6	5.1	5.8	4.4	5.2	-21.2	-20.2	-81.5	-79.5	18.4	19.6
2019/3/13	3.7	4.6	5.1	6.2	4.4	5.3	-20.9	-20.2	-81.5	-79.6	18.7	19.9
2019/3/14	3.7	4.6	5.4	6.2	4.4	5.2	-20.9	-19.9	-81.5	-79.2	18.8	20.1
2019/3/15	3.7	4.6	5.4	6.2	4.4	5.2	-20.4	-19.7	-81.5	-79.3	18.9	19.9
2019/3/16	4.2	4.6	5.7	6.2	4.4	4.9	-20.1	-19.8	-81.5	-79.7	18.7	19.1
2019/3/17	4.2	4.6	5.7	6.2	4.4	4.9	-20.1	-19.8	-81.5	-79.3	18.8	18.8
2019/3/18	3.7	4.7	5.5	6.2	4.4	5.3	-20.8	-19.8	-81.5	-79.6	18.8	20.1
2019/3/19	3.7	4.6	5.6	6.2	4.4	5.3	-20.9	-20.3	-81.5	-79.6	19.1	20.7
2019/3/20	3.7	4.6	5.6	6.2	4.3	5.2	-20.9	-20.4	-81.5	-79.6	19.8	21
2019/3/21	3.7	4.7	5.7	6.2	4.4	5.3	-20.9	-19.8	-81.5	-79.1	19.9	21.2
2019/3/22	3.7	4.7	5.7	6.2	4.3	5.2	-20.4	-18.8	-81.5	-79.6	19.9	21.2
2019/3/23	4.1	4.7	5.8	6.2	4.3	4.9	-21.2	-19.8	-81.5	-79.6	19.7	20.3
2019/3/24	4.2	4.6	5.8	6.2	4.3	4.9	-21.4	-20.9	-81.5	-79.7	19.5	19.7
2019/3/25	3.8	4.7	5.5	6.2	4.3	5.1	-21.3	-20.7	-81.5	-79.6	19.5	20.6
2019/3/26	3.8	4.6	5.4	6.2	4.2	5.1	-21.2	-19.1	-81.5	-79.2	19.6	21
2019/3/27	3.7	4.6	5.7	6.3	4.3	5	-20.7	-19.8	-81.4	-79.6	19.7	21.1
2019/3/28	3.7	4.7	5.6	6.2	4.3	5.3	-20.8	-19.6	-81.5	-79.4	19.5	21.4
2019/3/29	3.7	4.7	5.6	6.2	4.3	5.3	-21.2	-20.6	-81.5	-79.7	19.1	20.3
2019/3/30	4.2	4.7	5.7	6.2	4.4	4.9	-21.2	-20.9	-81.5	-79.7	18.8	19.4
2019/3/31	4.2	4.7	5.7	6.2	4.4	4.9	-21.2	-20.8	-81.5	-79.6	19.4	19.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 108
 藥學部 溫雅如 108
 師(三)級藥師 0401