

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年03月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max
2019/3/1	4.3	5.1	4.4	4.9	20.6	20.9
2019/3/2	4.3	5.1	4.4	4.9	20.3	20.8
2019/3/3	4.3	5.1	4.4	4.9	19.9	20.3
2019/3/4	4.2	5	4.4	5.1	19.6	20.3
2019/3/5	4.2	5.1	4.4	6.1	19.9	20.7
2019/3/6	4.3	5.2	4.4	4.9	20.2	20.9
2019/3/7	4.4	5.2	4.4	5.8	20.3	20.7
2019/3/8	4.4	5.4	4.4	4.9	20.1	20.5
2019/3/9	4.3	5.1	4.4	4.9	20.1	20.4
2019/3/10	4.3	5.1	4.4	4.8	19.9	20.2
2019/3/11	4.4	5.2	4.4	4.9	19.9	20.9
2019/3/12	4.3	5.1	4.4	5.3	20.1	21.1
2019/3/13	4.3	5.2	4.7	5.5	20.4	21.2
2019/3/14	4.4	5.2	4.3	5.9	20.7	21.6
2019/3/15	4.4	5.2	4.3	4.9	20.9	21.3
2019/3/16	4.4	5.1	4.3	4.9	20.5	21.1
2019/3/17	4.4	5.1	4.4	5	20.4	20.6
2019/3/18	4.4	5.2	4.4	5	20.3	21.4
2019/3/19	4.3	5.1	4.3	5.1	20.9	21.8
2019/3/20	4.4	5.3	4.7	5.1	21.4	22.2
2019/3/21	4.3	5.1	4.6	5.1	21.7	22.4
2019/3/22	4.3	5.1	4.5	5.2	21.9	22.6
2019/3/23	4.3	4.9	4.5	4.9	21.5	22.3
2019/3/24	4.3	4.9	4.5	4.9	21.1	21.5
2019/3/25	4.3	5.1	4.4	5.2	20.9	21.9
2019/3/26	4.2	5.1	4.3	4.9	21.7	22
2019/3/27	4.3	5.2	4.4	5	21.3	22.2
2019/3/28	4.2	5.4	4.5	5.8	21.5	22.8
2019/3/29	4.2	4.9	4.7	5.1	21.6	22.4
2019/3/30	4.2	4.9	4.7	5.1	21.3	21.9
2019/3/31	4.2	4.9	4.7	5.1	21.1	21.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 溫度均在正常範圍內

藥學部 邱保祥 108 0401
師(三)級藥師

藥學部 溫雅如 108 0401
師(三)級藥師