

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 108年03月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8℃		15~25℃	
日期	Min	Max	Min	Max
2019/3/1	3.9	4.3	21.7	21.9
2019/3/2	3.9	4.3	21.7	21.8
2019/3/3	3.9	4.3	21.7	21.8
2019/3/4	3.9	4.3	21.5	21.9
2019/3/5	3.9	4.2	21.7	22.2
2019/3/6	3.4	4.2	21.9	22.2
2019/3/7	3.4	4.7	21.6	22.1
2019/3/8	3.7	4.1	21.3	21.7
2019/3/9	3.7	4	21.5	21.7
2019/3/10	3.7	4.1	21.3	21.5
2019/3/11	3.7	4.1	20.9	21.4
2019/3/12	3.7	4	21.1	21.7
2019/3/13	3.5	4.2	21.3	21.9
2019/3/14	3.7	4.1	21.7	22.3
2019/3/15	3.8	4.1	21.9	22.1
2019/3/16	3.8	4.1	21.5	21.9
2019/3/17	3.9	4.2	21.8	22
2019/3/18	3.9	4.2	21.9	22.2
2019/3/19	3.9	4.2	21.9	22.4
2019/3/20	3.9	4.2	22.1	22.6
2019/3/21	3.9	4.1	22.2	23.1
2019/3/22	3.9	4.2	22.6	23.1
2019/3/23	3.9	4.2	22.2	23
2019/3/24	3.9	4.2	21.9	22.2
2019/3/25	3.9	4.2	22.2	22.4
2019/3/26	3.9	4.2	22	22.3
2019/3/27	3.8	4.2	22.2	22.6
2019/3/28	3.8	4.1	22.3	22.8
2019/3/29	3.8	4.1	22.6	23.1
2019/3/30	3.8	4.2	22.8	23
2019/3/31	3.8	4.1	22.6	23.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 108
師(三)藥師 邱保祥 0401
藥學部 謝雅如 108
師(一)藥師 謝雅如 0401