

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年02月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/2/1	3.5	4.5	4.7	5.5	4.4	5.3	-20.4	-19.1	-81.5	-79.6	19.6	20.3
2019/2/2	3.9	4.5	4.8	5.4	4.4	4.9	-20.2	-19.9	-81.5	-79.4	19.4	19.6
2019/2/3	3.9	4.5	4.8	5.5	4.4	4.9	-20.2	-19.9	-81.5	-79.3	19.3	19.4
2019/2/4	3.9	4.5	4.8	5.5	4.4	4.9	-20.2	-19.9	-81.5	-79.6	19.2	19.4
2019/2/5	3.9	4.5	4.8	5.5	4.4	4.9	-20.2	-19.9	-81.5	-79.7	19.1	19.2
2019/2/6	4	4.5	4.8	5.4	4.4	4.9	-20.2	-19.9	-81.5	-79.6	19.1	19.7
2019/2/7	3.9	4.5	4.8	5.5	4.4	4.9	-20.2	-19.9	-81.5	-79.2	19.6	19.8
2019/2/8	3.4	4.5	4.7	5.5	4.4	5.1	-20.2	-19.8	-81.5	-79.2	19.7	20.8
2019/2/9	3.9	4.5	4.7	5.4	4.4	4.9	-20.1	-19.9	-81.5	-79.7	19.9	19.9
2019/2/10	3.9	4.5	4.7	5.6	4.4	4.9	-20.2	-20	-81.5	-79.6	19.8	19.9
2019/2/11	3.4	4.5	4.7	5.6	4.4	5	-20.4	-19.8	-81.5	-77.6	19.4	20.8
2019/2/12	3.6	4.5	4.7	5.7	4.4	5.1	-20.4	-19.8	-81.5	-79.4	19.2	20.8
2019/2/13	3.6	4.5	4.7	5.8	4.4	5.3	-20.2	-19.5	-81.5	-79.3	19.2	20.4
2019/2/14	3.6	4.5	4.8	5.7	4.4	5.1	-19.7	-19	-81.5	-79.2	18.9	21.2
2019/2/15	3.4	4.5	4.8	5.7	4.4	5.5	-19.7	-19.3	-81.5	-79.6	19.7	20.9
2019/2/16	3.9	4.5	4.8	5.6	4.4	4.9	-19.8	-19.6	-81.5	-79.6	19.7	20.1
2019/2/17	3.9	4.5	4.8	5.6	4.4	4.9	-19.8	-19.7	-81.5	-79.1	19.4	19.7
2019/2/18	3.4	4.5	4.8	5.9	4.4	5.2	-19.9	-19.1	-81.5	-79.6	19.4	20.9
2019/2/19	3.4	4.4	4.8	5.7	4.5	5.3	-19.9	-19.6	-81.5	-79.4	19.3	20.8
2019/2/20	3.6	4.5	4.8	5.7	4.4	5.2	-20.4	-19	-81.5	-79.5	19	20.4
2019/2/21	3.5	4.5	4.8	5.7	4.4	5.2	-20.4	-19.1	-81.5	-79.4	18.9	19.9
2019/2/22	3.5	4.4	4.7	5.7	4.4	5.2	-20.4	-19.3	-81.5	-79.2	18.8	20
2019/2/23	3.9	4.4	4.7	5.6	4.3	4.9	-20.6	-20.2	-81.5	-79.4	18.7	19.8
2019/2/24	3.9	4.4	4.7	5.6	4.3	4.9	-20.6	-20.4	-81.5	-79.6	18.6	18.8
2019/2/25	3.5	4.4	2.6	5.7	4.3	5.1	-20.7	-20.2	-81.5	-79.7	18.4	20.1
2019/2/26	3.5	4.4	2.4	4.3	4.4	5.1	-20.8	-19.8	-81.5	-79.6	18.9	20.2
2019/2/27	3.6	4.4	3.2	5.3	4.3	5.1	-20.4	-19.6	-81.5	-79.5	19.3	20.7
2019/2/28	3.9	4.4	4.5	5.3	4.3	4.9	-19.9	-19.7	-81.5	-79.3	19.2	19.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內
 藥學部邱保祥 108
 師(三)藥師 0304
 藥學部溫雅如 108
 師(三)藥師 0204