

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年02月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max
2019/2/1	4.7	5.8	4.7	5.3	21.2	21.5
2019/2/2	4.6	5.3	4.7	5.2	20.9	21.3
2019/2/3	4.6	5.3	4.6	5.2	20.6	20.9
2019/2/4	4.6	5.3	4.6	5.1	20.6	20.8
2019/2/5	4.6	5.3	4.6	5.1	20.6	20.7
2019/2/6	4.6	5.3	4.7	5.1	20.6	21.1
2019/2/7	4.6	5.3	4.7	5.1	21.1	21.3
2019/2/8	4.6	5.4	4.2	5.1	21.1	21.3
2019/2/9	4.6	5.3	4.2	4.7	21.1	21.3
2019/2/10	4.6	5.3	4.2	4.7	20.9	21.2
2019/2/11	4.6	5.9	3.9	4.7	20.9	21.3
2019/2/12	4.6	5.9	3.9	4.7	20.8	21.6
2019/2/13	4.5	5.5	4	4.7	21.1	21.7
2019/2/14	4.6	5.6	4.1	4.7	21.1	22.1
2019/2/15	4.7	5.4	4	4.7	21.5	22.3
2019/2/16	4.7	5.4	4	4.6	21.7	22.2
2019/2/17	4.7	5.4	4	4.5	21.2	21.7
2019/2/18	4.4	5.6	4	5.8	20.9	22.1
2019/2/19	4.4	5.2	4.4	5.2	21.5	22.2
2019/2/20	4.4	5.3	4.7	5.5	21.3	21.9
2019/2/21	4.3	5.1	4.6	5.3	21.1	21.8
2019/2/22	4.3	5.2	4.6	5.3	21.1	21.7
2019/2/23	4.5	5.2	4.7	5.2	20.8	21.4
2019/2/24	4.4	5.2	4.5	5.1	20.3	20.8
2019/2/25	4.4	5.2	4.4	5.2	19.9	21
2019/2/26	4.4	5.3	4.3	4.9	20.6	21.4
2019/2/27	4.3	5.3	4.4	5.1	20.9	21.7
2019/2/28	4.3	5.1	4.5	4.9	20.9	21.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 108
師(三)級藥師 邱保祥 204
藥學部 溫雅如 108
師(三)級藥師 溫雅如 204