

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年01月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/1/1	3.8	4.4	4.3	5.1	4.4	4.9	-21.2	-20.9	-81.5	-79.7	20.1	20.2
2019/1/2	3.3	4.5	4.3	5.2	4.2	4.9	-22.4	-19.4	-81.5	-79.7	19.9	20.8
2019/1/3	3.4	4.4	4.2	5.2	4	4.8	-22.4	-19.6	-81.5	-79.7	19.7	20.9
2019/1/4	3.3	4.5	4.2	5.1	4.3	4.9	-20.7	-20	-81.5	-79.4	19.3	20.6
2019/1/5	4.1	4.5	4.2	5.1	4.4	4.9	-20.7	-20.4	-81.5	-79.7	19.3	19.4
2019/1/6	4.1	4.5	4.2	5.1	4.4	4.9	-20.7	-20.4	-81.5	-79.5	19.3	19.7
2019/1/7	3.7	4.5	4.2	5.1	4.3	4.9	-20.7	-20.1	-81.5	-79.4	18.9	20.1
2019/1/8	3.7	4.5	4.3	5.1	4.4	5	-20.4	-19.6	-81.5	-79.3	18.6	20.1
2019/1/9	3.7	4.5	4.3	5.1	4.3	5.2	-20.9	-19.8	-81.5	-79.7	19.2	20.4
2019/1/10	3.8	4.6	4.2	5.3	4.4	5.1	-20.9	-20.2	-81.5	-79.7	19.2	20.6
2019/1/11	3.7	4.5	4.4	5.3	4.4	5.1	-20.9	-20.2	-81.5	-79.5	19.3	20.4
2019/1/12	4	4.5	4.4	5.3	4.5	4.9	-20.9	-20.4	-81.5	-79.4	19.2	19.6
2019/1/13	4	4.5	4.4	5.3	4.5	4.9	-20.9	-20.4	-81.5	-79.4	19.1	19.5
2019/1/14	3.7	4.6	4.4	5.4	4.4	5	-20.9	-19.1	-81.5	-79.3	19.4	21.4
2019/1/15	3.7	4.6	4.4	5.3	4.3	5.1	-19.9	-18.6	-81.5	-77.9	19.6	20.7
2019/1/16	3.5	4.6	4.4	5.6	4.4	5.1	-20.9	-19.7	-81.5	-79.6	19.7	20.9
2019/1/17	3.5	4.5	4.7	5.5	4.6	5.2	-20.9	-16.5	-81.5	-79.4	19.6	20.8
2019/1/18	3.6	4.6	4.7	5.5	4.6	5.1	-20.8	-19.2	-81.5	-79.2	19.3	20.6
2019/1/19	4.2	4.6	4.7	5.5	4.6	5.1	-20.6	-20.2	-81.5	-79.7	19.4	19.6
2019/1/20	4.2	4.6	4.7	5.5	4.7	5.1	-20.4	-19.7	-81.5	-79.6	19.6	23.7
2019/1/21	3.5	4.6	4.6	5.5	4.6	5.1	-20.3	-19.8	-81.5	-79.6	20.3	21.8
2019/1/22	3.6	4.7	4.6	5.4	4.6	5.2	-20.3	-19.3	-81.5	-79	19.7	21
2019/1/23	3.6	4.4	4.6	5.5	4.7	5.2	-20.4	-19.5	-81.5	-79.3	19.5	20.8
2019/1/24	3.6	4.4	4.7	5.6	4.5	5.2	-20.4	-19.6	-81.5	-79.6	19.4	20.8
2019/1/25	3.6	4.4	4.6	5.5	4.5	5.3	-20.4	-19.1	-81.5	-79.7	19.7	20.9
2019/1/26	3.9	4.4	4.6	5.4	4.6	5.1	-20.9	-20.4	-81.5	-79.4	19.2	19.9
2019/1/27	3.9	4.4	4.7	5.5	4.6	5.1	-20.9	-20.7	-81.5	-79.4	19	19.2
2019/1/28	3.4	4.4	4.6	5.5	4.6	5.3	-20.9	-20.4	-81.5	-79.7	19	20.4
2019/1/29	3.4	4.4	4.6	5.4	4.6	5.4	-20.9	-20	-81.6	-79	19.4	20.8
2019/1/30	3.3	4.4	4.6	5.4	4.7	5.6	-21.2	-20.6	-81.5	-79.1	19.6	20.9
2019/1/31	3.3	4.4	4.2	5.4	4.7	5.6	-21.2	-20.7	-81.5	-77.4	19.7	20.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥 108
師(三)級藥師 201

藥學部 溫雅如 108
師(二)級藥師 201