

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 107年12月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-293		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-292		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2018/12/1	4.9	5.2	4.3	5.3	4.7	4.9	-22.4	-21.9	-81.4	-79.8	20.4	20.7
2018/12/2	4.9	5.2	4.4	5.3	4.7	4.9	-22.4	-21.9	-81.4	-79.8	20.3	20.4
2018/12/3	4.1	5.3	4.2	5.8	4.3	5	-21.7	-20.8	-81.4	-79.8	20.2	20.7
2018/12/4	4	5.3	4.2	5.7	4.3	5	-21.7	-20.9	-81.4	-79.4	20.5	21.1
2018/12/5	3.8	5.3	4.2	5.6	4.3	4.9	-21.7	-20.3	-81.4	-79.6	20.4	20.7
2018/12/6	3.9	5.3	4.2	5.5	4.2	5.1	-20.6	-19.8	-81.5	-79.8	20.2	20.6
2018/12/7	3.9	5.3	4.3	5.3	4.1	4.8	-20.9	-20.4	-81.4	-79.8	20.1	20.3
2018/12/8	4.8	5.2	4.3	5.2	4.4	4.8	-20.9	-20.6	-81.5	-79.8	19.8	20.1
2018/12/9	4.8	5.2	4.3	5.2	4.4	4.8	-20.9	-20.5	-81.4	-79.8	19.8	19.9
2018/12/10	3.8	5.2	4.3	5.4	4.2	4.9	-20.9	-20.4	-81.4	-79.8	19.8	20.4
2018/12/11	4.5	5.2	4.4	5.4	4.1	4.9	-21.1	-19.8	-81.5	-79.3	20.1	20.7
2018/12/12	4.2	5.3	3.9	5.3	4	4.9	-21.2	-20.8	-81.4	-79.8	20.2	20.6
2018/12/13	4.2	5.3	3.9	5.3	4.1	5	-21.2	-20.7	-81.4	-79.6	20.2	20.9
2018/12/14	4.2	5.1	4.4	5.3	4.1	5	-21.2	-20.7	-81.4	-79.8	20.5	21.1
2018/12/15	4.5	5.1	4.4	5.3	4.5	4.9	-21.2	-20.8	-81.5	-79.8	20.2	20.8
2018/12/16	4.5	5	4	5.4	4.5	4.9	-21.2	-20.8	-81.5	-79.8	19.7	20.2
2018/12/17	4.2	5.2	4	5.4	4	5	-21.6	-20.6	-81.5	-79.8	19.5	20.2
2018/12/18	4.2	5.2	4	5.3	4.1	5.6	-21.7	-21	-81.4	-79.3	19.8	20.4
2018/12/19	4.2	5.2	4.1	5.4	4.1	4.9	-21.7	-20.7	-81.4	-79.7	19.7	20.3
2018/12/20	4.2	5.2	4.3	5.4	4.2	5	-21.6	-21	-81.2	-79.6	19.8	20.3
2018/12/21	4.2	5.3	4.5	5.4	4.1	4.9	-21.4	-20.7	-81.2	-79.6	19.9	20.4
2018/12/22	4.3	5.3	4.5	5.4	4.2	4.9	-21.4	-20.8	-81.4	-79.6	20.1	20.3
2018/12/23	4.7	5.3	4.4	5.4	4.5	4.9	-21.2	-20.8	-81.2	-79.8	19.9	20.1
2018/12/24	4.1	5.2	4.5	5.5	4.1	4.9	-21.2	-19.6	-81.3	-79.6	19.8	20.4
2018/12/25	4.1	5.2	4.7	5.5	4	4.9	-21.4	-20.2	-81.3	-79.5	19.9	20.6
2018/12/26	4.1	5.2	4.7	5.5	4.1	4.9	-21.4	-20.9	-81.4	-79.3	20.1	20.6
2018/12/27	3.6	5.1	4.7	5.5	4.1	4.9	-21.4	-20.8	-81.3	-79.8	20.1	20.5
2018/12/28	3.6	4.9	4.6	5.5	4.2	4.9	-21.2	-20.8	-81.3	-79.8	19.9	20.3
2018/12/29	4.5	4.9	4.7	5.5	4.5	4.9	-21.4	-20.9	-81.4	-79.8	19.4	19.9
2018/12/30	4.5	5.1	4.7	5.5	4.5	4.9	-21.4	-20.9	-81.4	-79.6	19.2	19.6
2018/12/31	3.7	5	4.7	5.4	4.2	4.9	-21.2	-20.9	-81.4	-79.8	19.6	20.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部邱保祥 108
師(三)級藥師 邱保祥 1012
藥學部溫雅如 108
師(三)級藥師 溫雅如 1012