

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 107年12月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8℃		15~25℃	
日期	Min	Max	Min	Max
2018/12/1	3.7	4.3	20.5	21.1
2018/12/2	3.7	4.4	20.6	20.9
2018/12/3	3.7	4.3	20.5	21.1
2018/12/4	3.7	4.3	20.6	21.1
2018/12/5	3.4	4.3	20.9	21.3
2018/12/6	3.4	4	20.9	21.4
2018/12/7	3.3	4	20.8	21.3
2018/12/8	3.4	4	20.2	20.9
2018/12/9	3.4	4	20.1	20.6
2018/12/10	3.4	4.1	20.1	20.8
2018/12/11	3.4	4	20.3	21.1
2018/12/12	3.4	4.2	20.1	20.7
2018/12/13	3.7	4.3	20.3	20.7
2018/12/14	3.4	4.3	19.8	20.4
2018/12/15	3.4	4.2	19.8	20.2
2018/12/16	3.4	4.1	19.7	19.9
2018/12/17	3.4	4.2	19.3	20.7
2018/12/18	3.5	4.2	19.5	19.9
2018/12/19	3.5	4.2	19.6	20.1
2018/12/20	3.4	4.4	19.7	20.5
2018/12/21	3.4	4.2	19.9	20.3
2018/12/22	3.3	5.7	20.2	20.5
2018/12/23	3.3	4.1	19.7	20.3
2018/12/24	3.3	4.1	19.7	20.3
2018/12/25	3.2	4.1	20.2	22.6
2018/12/26	2.5	4.1	22.6	23
2018/12/27	2.5	4.7	22.6	22.9
2018/12/28	4	4.3	22.1	22.6
2018/12/29	4	4.3	21.8	22
2018/12/30	4.1	4.3	21.8	22.1
2018/12/31	4	4.3	21.8	22

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

 藥學部 邱保祥 108
 藥學部 溫雅如 108