

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 10 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/10/1	4.2	4.7	4	4.4	3.3	4.2	-20.2	-19.9	NA	NA	19.7	19.9
2020/10/2	4.3	4.7	4	4.4	3.3	4.2	-20.2	-19.9	NA	NA	19.5	19.7
2020/10/3	4.2	4.7	4	4.4	3.3	4.2	-20.2	-19.9	NA	NA	19.4	19.5
2020/10/4	4.2	4.7	4	4.4	3.3	4.2	-20.2	-19.9	NA	NA	19.4	19.4
2020/10/5	3.9	5.7	3.8	4.8	3.3	5.9	-19.9	-18.7	NA	NA	19.4	20.1
2020/10/6	4.9	5.7	3.7	4.6	3.3	5.3	-20.4	-19.8	NA	NA	19.5	20.2
2020/10/7	4.9	5.8	3.7	4.7	3.3	5.3	-20.4	-19.9	NA	NA	19.6	20.4
2020/10/8	5.1	5.9	3.8	4.7	3.4	6.1	-20.4	-19.9	NA	NA	19.6	20.3
2020/10/9	5.4	5.9	3.9	4.4	3.5	4.2	-20.2	-19.9	NA	NA	19.6	19.8
2020/10/10	5.4	5.9	3.9	4.4	3.5	4.2	-20.2	-19.9	NA	NA	19.4	19.6
2020/10/11	5.4	5.9	3.9	4.4	3.5	4.2	-20.2	-19.9	NA	NA	19.4	19.7
2020/10/12	5.1	5.9	3.6	4.5	3.5	5.9	-20.2	-19.8	NA	NA	19.3	19.9
2020/10/13	5.1	5.7	3.6	4.6	3.5	5.5	-20.1	-18.9	NA	NA	19.5	20.2
2020/10/14	5.1	5.8	3.7	4.4	3.6	5.4	-20.4	-16.8	NA	NA	19.4	20.1
2020/10/15	5.2	5.9	3.7	4.6	3.6	4.7	-19.7	-18.7	NA	NA	19.5	20.1
2020/10/16	5.1	6.1	3.6	4.4	3.6	4.9	-20.2	-17.6	NA	NA	19.6	20.9
2020/10/17	5.7	6.1	3.8	4.4	3.6	4.2	-20.3	-20	NA	NA	19.5	19.7
2020/10/18	5.7	6.1	3.8	4.4	3.6	4.2	-20.4	-20.1	NA	NA	19.3	19.5
2020/10/19	5.3	6.1	3.5	4.7	3.6	4.7	-20.4	-18.1	NA	NA	19.3	20.2
2020/10/20	5.3	6.1	3.7	4.4	3.5	4.7	-20.2	-18.7	NA	NA	19.4	19.9
2020/10/21	5.3	6.2	3.9	5.1	3.3	4.3	-19.8	-18.4	NA	NA	19.4	20.2
2020/10/22	5.3	6	4.2	4.9	3.3	5.1	-19.7	-18.7	NA	NA	19.5	20.3
2020/10/23	5.3	6	4.3	4.9	3.6	5.4	-20.6	-19	NA	NA	19.4	20.3
2020/10/24	5.3	5.8	4.4	4.8	3.7	4.2	-20.4	-20.2	NA	NA	19.2	19.5
2020/10/25	5.3	5.8	4.4	4.9	3.7	4.2	-20.4	-20.2	NA	NA	19.1	19.2
2020/10/26	5.3	6	4.3	4.9	3.7	6.1	-20.4	-19.7	NA	NA	19	19.8
2020/10/27	5.2	6.3	4.3	4.9	3.7	5.7	-19.9	-19.3	NA	NA	19.2	19.8
2020/10/28	5.2	6	4.3	4.8	3.7	5.6	-20.4	-19.6	NA	NA	19.3	19.9
2020/10/29	5.3	5.8	4.3	4.8	3.7	5.6	-20.4	-19.3	NA	NA	19.3	20.1
2020/10/30	5.3	5.9	4.3	4.8	3.7	5.3	-19.6	-19.1	NA	NA	19.4	20.5
2020/10/31	5.3	5.7	4.4	4.8	3.7	4.2	-19.6	-19.5	NA	NA	19.3	19.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥 109/102
師(三)級藥師

藥學部 四雅如 109/102
師(一)級藥師