

# 臺北榮民總醫院藥學部

## 臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 109 年 10 月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2020/10/1	4.9	5.3	19.2	19.4
2020/10/2	4.9	5.3	18.4	19.3
2020/10/3	4.8	5.3	18.2	18.6
2020/10/4	4.8	5.2	18.2	18.9
2020/10/5	4.7	5.4	18.4	19.1
2020/10/6	4.5	5.3	18.8	19.4
2020/10/7	4.7	5.3	19.2	19.7
2020/10/8	4.7	5.4	19.2	19.7
2020/10/9	4.7	5.3	19.2	19.4
2020/10/10	4.7	5.3	18.4	19.2
2020/10/11	4.7	5.3	18.2	18.8
2020/10/12	4.7	5.3	18	18.8
2020/10/13	4.7	5.2	18.4	18.8
2020/10/14	4.6	5.3	18.6	18.9
2020/10/15	4.9	5.3	18.8	19.3
2020/10/16	4.9	5.3	18.6	19.2
2020/10/17	4.9	5.3	19	19.4
2020/10/18	4.9	5.3	18.9	19.2
2020/10/19	4.8	5.4	18.8	19.4
2020/10/20	4.9	5.4	18.9	19.4
2020/10/21	4.9	5.4	19.3	19.8
2020/10/22	4.9	5.4	19.8	20.4
2020/10/23	5.1	5.4	19.9	20.6
2020/10/24	4.9	5.5	19.8	20.4
2020/10/25	4.9	5.4	19.5	20.2
2020/10/26	4.7	5.3	19.7	20.4
2020/10/27	4.6	5.1	19.8	20.3
2020/10/28	4.5	5.1	19.7	20.5
2020/10/29	4.8	5.2	19.8	20.3
2020/10/30	4.8	5.2	20	20.5
2020/10/31	4.8	5.1	19.6	20.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 1091102  
藥學部 溫雅如 1091102

溫度均在正常範圍內