

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 11 月

|            | 一號冰箱      |     | 二號冰箱     |     | 三號冰箱      |     | 負20度冰箱    |       | 負80度冰箱    |     | 室溫1       |      |
|------------|-----------|-----|----------|-----|-----------|-----|-----------|-------|-----------|-----|-----------|------|
| 序號         | 0520-362  |     | 0520-292 |     | 1504-0305 |     | 1405-0193 |       | 1215-0006 |     | 1213-0007 |      |
| 備用         | 0850-0835 |     | 0520-293 |     | 1504-0339 |     | 1213-0010 |       | 1330-0064 |     | 1405-0194 |      |
| 範圍         | 2~8°C     |     | 2~8°C    |     | 2~8°C     |     | 負15~負25°C |       | ≤負70°C    |     | 15~25°C   |      |
| 日期         | Min       | Max | Min      | Max | Min       | Max | Min       | Max   | Min       | Max | Min       | Max  |
| 2020/11/1  | 5.3       | 5.7 | 4.4      | 4.8 | 3.7       | 4.2 | -19.6     | -19.6 | NA        | NA  | 19.2      | 19.3 |
| 2020/11/2  | 5.3       | 6   | 4.2      | 4.8 | 3.8       | 5.1 | -19.3     | -17.9 | NA        | NA  | 19.4      | 19.8 |
| 2020/11/3  | 5.2       | 5.9 | 4.1      | 4.8 | 3.7       | 4.9 | -20.3     | -19   | NA        | NA  | 19.2      | 19.9 |
| 2020/11/4  | 5.3       | 5.9 | 4.2      | 4.8 | 3.7       | 5.4 | -20.2     | -19.4 | NA        | NA  | 19.4      | 20.5 |
| 2020/11/5  | 5.4       | 5.9 | 4.2      | 4.8 | 3.9       | 5.4 | -19.6     | -18.8 | NA        | NA  | 19.9      | 20.6 |
| 2020/11/6  | 5.3       | 5.9 | 4.2      | 4.8 | 3.9       | 5.8 | -19.4     | -16.6 | NA        | NA  | 20.1      | 20.8 |
| 2020/11/7  | 5.4       | 5.9 | 4.3      | 4.9 | 3.9       | 4.4 | -19.3     | -18.7 | NA        | NA  | 20.1      | 20.3 |
| 2020/11/8  | 5.4       | 5.8 | 4.3      | 4.9 | 3.9       | 4.4 | -19.3     | -19   | NA        | NA  | 19.9      | 20.1 |
| 2020/11/9  | 5         | 5.8 | 4.2      | 5   | 3.9       | 5.9 | -19.3     | -18.7 | NA        | NA  | 19.7      | 20.3 |
| 2020/11/10 | 4.9       | 5.9 | 4.3      | 4.8 | 3.8       | 5.8 | -19.7     | -18.7 | NA        | NA  | 19.8      | 20.4 |
| 2020/11/11 | 4.9       | 5.7 | 4.2      | 4.9 | 3.8       | 5.8 | -19.9     | -19.5 | NA        | NA  | 19.9      | 20.6 |
| 2020/11/12 | 4.8       | 6.5 | 4.2      | 4.9 | 3.7       | 5.2 | -19.9     | -19   | NA        | NA  | 19.9      | 23.2 |
| 2020/11/13 | 4.9       | 5.6 | 4.2      | 4.7 | 3.7       | 5.2 | -19.9     | -19   | NA        | NA  | 20.1      | 20.8 |
| 2020/11/14 | 4.9       | 5.6 | 4.3      | 4.7 | 3.7       | 4.3 | -19.9     | -19.7 | NA        | NA  | 20.1      | 20.3 |
| 2020/11/15 | 4.9       | 5.6 | 4.3      | 4.7 | 3.7       | 4.3 | -19.9     | -19.7 | NA        | NA  | 19.9      | 20.6 |
| 2020/11/16 | 4.9       | 6.1 | 4.2      | 4.9 | 3.8       | 5.3 | -19.9     | -18.7 | NA        | NA  | 19.9      | 20.7 |
| 2020/11/17 | 4.9       | 5.9 | 4.3      | 4.8 | 3.8       | 5.8 | -19.9     | -19.1 | NA        | NA  | 20.1      | 20.8 |
| 2020/11/18 | 5.2       | 5.9 | 4.3      | 4.8 | 3.5       | 4.7 | -19.9     | -19.7 | NA        | NA  | 20.1      | 20.9 |
| 2020/11/19 | 5.3       | 5.9 | 4.3      | 4.9 | 3.5       | 4.9 | -19.9     | -19   | NA        | NA  | 20.1      | 20.8 |
| 2020/11/20 | 5.1       | 6.1 | 4.3      | 4.8 | 3.9       | 4.8 | -19.3     | -18.7 | NA        | NA  | 20.1      | 20.8 |
| 2020/11/21 | 5.1       | 5.7 | 4.4      | 4.8 | 4         | 4.4 | -18.9     | -18.9 | NA        | NA  | 20.2      | 20.3 |
| 2020/11/22 | 5.1       | 5.7 | 4.4      | 4.8 | 3.9       | 4.4 | -19.1     | -18.8 | NA        | NA  | 20.1      | 20.2 |
| 2020/11/23 | 5.1       | 6.1 | 4.3      | 4.8 | 3.7       | 4.6 | -19.1     | -18.7 | NA        | NA  | 19.7      | 20.2 |
| 2020/11/24 | 5.3       | 6.1 | 4.3      | 4.8 | 3.9       | 4.6 | -19.1     | -18.7 | NA        | NA  | 19.5      | 20.1 |
| 2020/11/25 | 5.4       | 6.2 | 4.3      | 4.8 | 3.9       | 4.7 | -19.1     | -18.7 | NA        | NA  | 19.6      | 20.2 |
| 2020/11/26 | 5.5       | 6.1 | 4.2      | 4.8 | 4         | 4.7 | -19.1     | -18.7 | NA        | NA  | 19.5      | 20.4 |
| 2020/11/27 | 5.6       | 6.2 | 4.2      | 4.8 | 4.2       | 4.9 | -19.1     | -17.4 | NA        | NA  | 19.7      | 20.3 |
| 2020/11/28 | 5.6       | 6   | 4.3      | 4.8 | 4.3       | 4.7 | -19.1     | -18.8 | NA        | NA  | 19.4      | 19.7 |
| 2020/11/29 | 5.6       | 6   | 4.3      | 4.7 | 4.3       | 4.7 | -19.1     | -19   | NA        | NA  | 19.2      | 19.4 |
| 2020/11/30 | 5.2       | 6.2 | 4.3      | 4.8 | 4.3       | 5.1 | -19.1     | -18.5 | NA        | NA  | 19.2      | 19.9 |

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥  
師(三)級藥師 20109

藥學部 溫雅如  
師(三)級藥師 20109