

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 11 月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2020/11/1	3.9	4.6	4.7	5.6	20.8	21.1
2020/11/2	3.6	4.4	4.7	5.7	20.7	20.9
2020/11/3	3.6	4.4	4.8	5.7	20.9	21.2
2020/11/4	3.5	4.6	4.6	5.7	21.2	21.4
2020/11/5	3.5	4.4	4.7	5.8	21.3	21.7
2020/11/6	3.3	4.5	4.7	5.7	21.6	21.9
2020/11/7	3.4	4.3	4.7	5.6	21.7	21.9
2020/11/8	3.4	4.3	4.7	5.6	21.4	21.7
2020/11/9	3.3	4.6	4.7	5.6	21.3	21.4
2020/11/10	3.4	4.5	4.6	5.8	21.3	21.6
2020/11/11	3.5	4.5	4.8	5.7	21.3	21.6
2020/11/12	3.6	4.5	4.7	5.7	21.4	21.7
2020/11/13	3.7	4.6	4.7	5.6	21.6	21.9
2020/11/14	3.7	4.5	4.8	5.6	21.7	21.8
2020/11/15	3.7	4.5	4.8	5.6	21.4	21.7
2020/11/16	3.6	5.1	4.8	5.7	21.3	21.8
2020/11/17	3.8	4.8	4.3	6.2	21.8	21.9
2020/11/18	3.9	5.3	3.9	5.5	21.7	21.9
2020/11/19	3.9	5.2	4.5	5.6	21.8	22.1
2020/11/20	3.8	5.1	4.7	5.5	21.9	22.2
2020/11/21	3.8	4.6	4.7	5.4	21.9	22.2
2020/11/22	3.8	4.6	4.7	5.4	21.7	21.9
2020/11/23	3.8	4.7	4.5	6.1	21.5	21.7
2020/11/24	3.9	4.7	4.3	5.4	21.5	21.7
2020/11/25	3.9	4.6	4.4	5.6	21.4	21.7
2020/11/26	3.9	6.9	4.7	6.3	21.4	21.7
2020/11/27	4.1	5.2	4.8	5.7	21.5	21.7
2020/11/28	4.2	4.9	4.8	5.5	21.3	21.7
2020/11/29	4.2	4.9	4.8	5.6	21.1	21.3
2020/11/30	3.9	5.7	4.5	5.8	20.9	21.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 溫度均在正常範圍內
藥學部 邱保祥
師(三)級藥師 109/12/01
藥學部 溫雅如
師(三)級藥師 109/12/01