

# 臺北榮民總醫院藥學部

## 臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 109 年 11 月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2020/11/1	4.8	5.1	19.6	20.4
2020/11/2	4.7	5.2	19.9	20.4
2020/11/3	4.6	4.9	20.1	20.6
2020/11/4	4.7	5.1	19.9	20.4
2020/11/5	4.7	5.1	19.8	20.9
2020/11/6	4.7	4.9	19.7	20.7
2020/11/7	4.7	4.9	19.7	20.6
2020/11/8	4.7	5.1	19.9	20.5
2020/11/9	4.3	6.3	19.8	20.6
2020/11/10	4.4	4.9	19.4	20.2
2020/11/11	4.4	5	19.6	20.2
2020/11/12	4.5	5	19.7	20.7
2020/11/13	4.6	4.9	20.1	20.6
2020/11/14	4.6	5	19.9	20.3
2020/11/15	4.6	5.1	19.4	20.3
2020/11/16	4.7	5.1	19.4	20.1
2020/11/17	4.7	5.1	19.6	20.3
2020/11/18	4.4	5.1	19.6	19.8
2020/11/19	4.7	5.2	19.5	20.3
2020/11/20	4.7	5.1	19.4	20.1
2020/11/21	4.7	5.1	19.4	19.7
2020/11/22	4.7	5.1	19.4	20.2
2020/11/23	4.7	5.1	19.8	20.4
2020/11/24	4.7	5.1	19.8	20.4
2020/11/25	4.4	5.1	19.7	20.7
2020/11/26	4.7	5.1	19.7	20.7
2020/11/27	4.7	5.2	19.9	20.9
2020/11/28	4.9	5.2	19.8	20.2
2020/11/29	4.9	5.2	19.3	19.8
2020/11/30	4.9	5.2	19.1	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內  
藥學部 邱保祥 109/201  
師(三)級藥師  
藥學部 邱雅如 109/201  
師(二)級藥師