

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 12 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2020/12/1	5.2	6.1	4.1	5.2	4.1	5.3	-19.4	-18.9	NA	NA	20.1	20.4
2020/12/2	5.4	6.2	4.2	5	3.9	5.2	-19.8	-18.5	NA	NA	20.1	20.7
2020/12/3	5.3	6.2	4.3	4.8	3.9	5.8	-19.8	-17.1	NA	NA	20.1	20.4
2020/12/4	5.3	6.2	4.3	5.8	3.9	5.7	-19.6	-18.9	NA	NA	20.1	20.5
2020/12/5	5.7	6.2	4.3	4.7	4.1	4.4	-19.6	-19.6	NA	NA	20.1	20.2
2020/12/6	5.7	6.2	4.3	4.7	4.1	4.4	-19.6	-19.6	NA	NA	19.9	20.1
2020/12/7	5.5	6.4	4.3	4.7	4.1	5.4	-20.4	-19.6	NA	NA	19.8	20.3
2020/12/8	5.5	6.3	4.3	4.7	4	6.7	-20.6	-19.8	NA	NA	19.9	20.5
2020/12/9	5.4	6.1	4.3	5.1	3.9	4.7	-20.6	-19.8	NA	NA	20.1	20.5
2020/12/10	5.4	6.4	4.4	4.8	3.9	4.9	-20.1	-19	NA	NA	20.1	20.6
2020/12/11	5.4	6.1	4.4	5.1	3.9	5.1	-19.9	-19.1	NA	NA	20.1	20.6
2020/12/12	5.7	6.1	4.4	4.8	3.9	4.4	-19.9	-19.3	NA	NA	20.2	20.4
2020/12/13	5.7	6.1	4.4	4.7	3.9	4.4	-19.9	-19.9	NA	NA	20.1	20.2
2020/12/14	5.3	6.1	4.3	4.7	3.9	5	-19.9	-19.1	NA	NA	19.9	20.3
2020/12/15	5.2	6.1	4.3	4.7	4.1	5.1	-19.8	-19.2	NA	NA	19.9	20.4
2020/12/16	5.1	6.1	4.3	4.8	3.9	5.4	-19.9	-18.8	NA	NA	20.1	20.5
2020/12/17	5.4	6	4.3	4.9	3.6	4.8	-19.9	-19.4	NA	NA	20.1	20.8
2020/12/18	5.4	6.1	4.3	4.9	3.5	5.1	-20.4	-19.6	NA	NA	20.4	20.8
2020/12/19	5.5	6	4.3	4.8	3.8	4.3	-20.4	-20.2	NA	NA	19.9	20.4
2020/12/20	5.5	6	4.3	4.8	3.8	4.3	-20.4	-20.2	NA	NA	19.6	19.9
2020/12/21	5.3	6	4.3	4.8	3.7	4.5	-20.4	-20.1	NA	NA	19.4	19.7
2020/12/22	4.9	6	4.3	4.8	3.7	4.6	-20.4	-20	NA	NA	19.5	19.9
2020/12/23	5.1	5.9	4.4	4.8	3.7	4.4	-20.4	-20	NA	NA	19.6	20.2
2020/12/24	5.1	5.9	4.4	4.8	3.7	5.2	-20.4	-19.6	NA	NA	20.1	20.7
2020/12/25	5.2	5.9	4.4	4.8	3.4	4.5	-20.1	-19.6	NA	NA	20.2	20.6
2020/12/26	5.3	5.9	4.4	4.8	3.7	4.3	-20.1	-19.8	NA	NA	20.2	20.3
2020/12/27	5.3	5.9	4.4	4.8	3.7	4.3	-20.1	-19.9	NA	NA	20.1	20.2
2020/12/28	5.1	5.9	4.4	4.9	3.3	4.3	-20.1	-19.6	NA	NA	20.1	20.5
2020/12/29	5.3	5.9	4.4	4.9	3.3	4.4	-19.8	-19.6	NA	NA	20.2	20.7
2020/12/30	5.2	5.9	4.4	5	3.4	4.3	-20.4	-19.6	NA	NA	19.8	20.5
2020/12/31	5.2	5.9	4.4	5.1	3.4	4.3	-20.4	-19.7	NA	NA	19.3	19.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥
師(三)級藥師 110 0104

藥學部 溫雅如
師(二)級藥師 110 0106