

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 109 年 12 月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2020/12/1	4.1	4.9	4.7	5.7	21.1	21.3
2020/12/2	3.8	4.9	4.7	5.6	21.1	21.3
2020/12/3	3.8	4.9	3.7	5.9	21.1	21.3
2020/12/4	3.8	4.9	3.9	5.4	21.1	21.3
2020/12/5	3.9	4.8	4.6	5.4	21.1	21.2
2020/12/6	3.9	4.8	4.6	5.4	20.8	21.1
2020/12/7	3.8	4.8	4.7	5.6	20.7	21.1
2020/12/8	3.7	4.8	4.7	6.2	20.9	21.1
2020/12/9	3.6	6	4.7	6.1	20.9	21.3
2020/12/10	3.6	4.7	4.7	5.8	21.1	21.3
2020/12/11	3.7	4.7	4.7	5.6	21.1	21.3
2020/12/12	3.7	4.7	4.7	5.4	21.1	21.3
2020/12/13	3.7	4.7	4.7	5.4	20.8	21.1
2020/12/14	3.7	4.7	4.7	5.8	20.7	20.9
2020/12/15	3.6	5	4.4	6.1	20.7	21.1
2020/12/16	3.5	4.9	4.3	5.7	20.9	21.1
2020/12/17	3.5	4.7	4.6	5.8	20.9	21.3
2020/12/18	3.7	4.8	5	6.1	21.1	21.4
2020/12/19	3.8	4.8	5.4	6.1	20.4	21.1
2020/12/20	3.8	4.7	5.4	6.1	20.1	20.4
2020/12/21	3.7	5.4	5.5	6.2	19.9	20.2
2020/12/22	3.7	4.7	5.6	6.3	19.9	20.3
2020/12/23	3.7	4.7	5.5	6.2	20.2	20.6
2020/12/24	3.5	4.8	5	6.3	20.6	20.9
2020/12/25	3.7	4.7	5.1	5.9	20.7	20.9
2020/12/26	3.7	4.7	5.1	5.9	20.8	20.9
2020/12/27	3.7	4.7	5.1	5.9	20.6	20.7
2020/12/28	3.7	5.1	5	6.2	20.5	20.8
2020/12/29	3.9	5.3	5.3	6.4	20.7	21.1
2020/12/30	3.9	6.7	5.5	6.3	20.5	21.1
2020/12/31	3.9	5.1	5.5	6.3	19.9	20.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 110 0104  
藥學部 溫雅如 110 0106