

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 01 月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2021/1/1	3.9	4.8	5.7	6.3	19.7	20.3
2021/1/2	3.9	4.8	5.7	6.3	20.3	20.8
2021/1/3	3.9	4.8	5.7	6.2	20.1	20.5
2021/1/4	3.8	4.8	5.7	6.3	20.1	20.2
2021/1/5	3.9	5.1	5.6	6.3	20.2	20.5
2021/1/6	3.9	5.3	5.2	6.2	20.3	20.6
2021/1/7	4.9	5.5	5.1	6.1	20.3	20.6
2021/1/8	4.8	5.4	5.2	6	19.6	20.3
2021/1/9	4.8	5.3	5.2	6	19.1	19.6
2021/1/10	4.9	5.3	5.2	6	18.9	19.1
2021/1/11	4.9	5.4	5.2	6.1	18.9	19.3
2021/1/12	4.8	5.6	5.3	6.1	19	19.2
2021/1/13	4.9	5.4	5.3	6.3	18.8	19.2
2021/1/14	4.9	5.5	5.7	6.3	19.2	19.4
2021/1/15	4.9	5.4	5.7	6.3	19.4	19.8
2021/1/16	4.9	5.3	5.7	6.3	19.8	19.8
2021/1/17	4.9	5.3	5.7	6.3	19.5	19.8
2021/1/18	4.9	5.4	5.7	6.3	19.4	19.5
2021/1/19	4.8	5.4	5.6	6.2	19.4	19.8
2021/1/20	4.7	5.4	5.6	6.1	19.8	20.3
2021/1/21	4.7	5.8	5.6	6.2	20.3	20.7
2021/1/22	4.7	5.4	5.6	6.1	20.6	20.9
2021/1/23	4.9	5.3	5.6	6.1	20.8	20.9
2021/1/24	4.8	5.3	5.6	6.1	20.5	20.8
2021/1/25	4.8	5.7	5.6	6.3	20.4	20.7
2021/1/26	4.8	5.4	5.7	6.2	20.6	20.9
2021/1/27	4.8	5.4	5.4	7.1	20.8	21.1
2021/1/28	4.8	5.9	5.3	6.1	20.8	20.9
2021/1/29	4.9	6.6	5.5	6.4	20.4	20.8
2021/1/30	5.3	5.7	5.6	6.1	20.3	20.4
2021/1/31	5.3	5.7	5.6	6.2	20.1	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 110
師(三)級藥師 2201 藥學部 溫雅如 110
師(三)級藥師 2204