

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年02月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007						
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/2/1	4.6	5.7	4.3	4.9	4	4.7	-22.2	-18.2	NA	NA	19	19.7
2021/2/2	4.4	6.1	4.4	4.9	4.1	4.7	-22.2	-18.7	NA	NA	18.9	19.6
2021/2/3	4.4	6.1	4.4	4.9	4	4.7	-20.9	-18.9	NA	NA	18.8	21
2021/2/4	5.3	5.9	4.4	5.6	3.9	4.7	-20.6	-19.6	NA	NA	19.1	19.9
2021/2/5	5.1	6	4.3	5.3	4	5.1	-20.6	-19.6	NA	NA	19.2	19.8
2021/2/6	5.1	5.6	4.3	4.8	4.4	4.9	-20.6	-20.3	NA	NA	19.2	19.3
2021/2/7	5.1	5.6	4.4	4.8	4.4	4.8	-20.7	-20.3	NA	NA	19	19.2
2021/2/8	4.9	5.7	4.3	5.1	3.9	4.8	-20.6	-19.9	NA	NA	18.9	19.6
2021/2/9	4.9	5.7	4.3	5.1	3.9	4.8	-20.4	-19.9	NA	NA	18.9	19.5
2021/2/10	5.3	5.7	4.3	4.8	4.5	4.9	-20.4	-20.2	NA	NA	19	19.2
2021/2/11	5.3	5.7	4.3	4.8	4.5	4.9	-20.4	-20.2	NA	NA	18.8	19
2021/2/12	5.3	5.7	4.3	4.8	4.4	4.9	-20.4	-20.2	NA	NA	18.8	18.8
2021/2/13	5.3	5.7	4.3	4.8	4.5	4.9	-20.4	-20.2	NA	NA	18.7	18.8
2021/2/14	5.3	5.7	4.3	4.8	4.4	4.9	-20.4	-20.2	NA	NA	18.7	18.7
2021/2/15	5.3	5.7	4.3	4.8	4.4	4.9	-20.6	-20.3	NA	NA	18.6	18.7
2021/2/16	4.9	5.7	4.3	4.9	4	4.9	-20.6	-19.9	NA	NA	18.4	18.9
2021/2/17	5	5.7	4.3	5.1	4	4.9	-20.2	-19.6	NA	NA	18.6	19.1
2021/2/18	4.9	5.8	4.3	4.8	4.1	4.9	-19.9	-19.3	NA	NA	18.4	19
2021/2/19	4.8	5.9	4.3	4.8	4.1	4.9	-20.8	-19.7	NA	NA	18.2	19.3
2021/2/20	4.8	5.9	4.3	4.8	4.2	4.8	-20.9	-20.1	NA	NA	18	19.3
2021/2/21	5.3	5.9	4.3	4.7	4.4	4.8	-20.7	-20.4	NA	NA	17.9	18.1
2021/2/22	5	5.9	4.3	4.8	4.1	4.7	-20.7	-20.2	NA	NA	17.8	18.7
2021/2/23	5.1	5.9	4.3	4.8	4	4.8	-20.7	-20	NA	NA	17.9	18.6
2021/2/24	4.8	6	4.3	4.7	3.9	4.9	-20.8	-20.2	NA	NA	18	18.7
2021/2/25	4.7	6	4.3	4.7	3.9	4.9	-20.9	-20.4	NA	NA	18.2	18.9
2021/2/26	4.7	5.7	4.3	4.8	3.9	5	-20.9	-20.3	NA	NA	18.4	19.1
2021/2/27	5.2	5.7	4.3	4.8	4.7	5.1	-20.6	-20.3	NA	NA	18.4	18.6
2021/2/28	5.2	5.7	4.3	4.8	4.7	5.1	-20.6	-20.4	NA	NA	18.3	18.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 110 0302
師(三)級藥師
藥學部 溫雅如 110 302 664
師(二)級藥師