

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年03月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/3/1	5.2	5.7	4.3	4.8	4.7	5.1	-20.7	-20.5	NA	NA	18.2	18.3
2021/3/2	4.9	5.8	4.2	4.8	4.1	5.2	-20.2	-19.8	NA	NA	18.3	18.7
2021/3/3	5.1	5.8	4.2	4.8	4	5.2	-20.8	-20	NA	NA	18.2	19.1
2021/3/4	5.2	5.8	4.2	5.1	3.9	5.3	-20.8	-19.9	NA	NA	18.4	19.1
2021/3/5	4.9	5.8	4.4	4.9	4	4.9	-20.4	-19.9	NA	NA	18.4	19.2
2021/3/6	5.1	5.7	4.4	4.9	4.6	5	-20.4	-20.2	NA	NA	18.4	18.7
2021/3/7	5.1	5.7	4.4	4.9	4.7	5	-20.4	-20.2	NA	NA	18.4	18.4
2021/3/8	4.9	5.7	4.4	5.1	4	5	-20.8	-20.1	NA	NA	18.3	18.9
2021/3/9	4.9	6	4.2	5.2	4	5.1	-20.8	-20.2	NA	NA	18.4	19.1
2021/3/10	5.1	6	4.3	4.8	4	5.1	-20.7	-20.2	NA	NA	18.4	19.2
2021/3/11	5.1	5.8	4.1	5.3	3.9	4.9	-20.7	-19.8	NA	NA	18.4	19.2
2021/3/12	5.2	5.8	4.2	5.1	4.1	5	-20.4	-19.9	NA	NA	18.6	19.3
2021/3/13	5.2	5.7	4.3	4.8	4.5	5	-20.1	-19.9	NA	NA	18.4	18.6
2021/3/14	5.2	5.7	4.3	4.8	4.5	5	-20.2	-19.9	NA	NA	18.3	18.4
2021/3/15	4.8	5.9	4.2	5	4.1	5	-20.9	-20	NA	NA	18.2	19.2
2021/3/16	4.7	5.7	4.2	4.8	4.2	5	-20.9	-19.9	NA	NA	18.4	19.1
2021/3/17	4.7	5.7	4	5	4.2	5.1	-20.4	-19.9	NA	NA	18.4	19.2
2021/3/18	4.7	5.6	4.2	4.7	4	5.1	-20.6	-20	NA	NA	18.6	19.3
2021/3/19	4.7	5.6	4.2	4.8	4	4.9	-20.7	-19.5	NA	NA	18.7	19.2
2021/3/20	5.1	5.6	4.3	4.8	4.5	4.9	-19.7	-19.7	NA	NA	18.6	18.7
2021/3/21	5.1	5.6	4.3	4.8	4.5	5	-19.7	-19.6	NA	NA	18.2	18.6
2021/3/22	4.6	5.7	4.3	4.9	3.9	4.9	-19.8	-19.6	NA	NA	18.1	18.9
2021/3/23	4.6	5.6	4.4	5	3.8	4.8	-20.9	-19.7	NA	NA	18.1	18.8
2021/3/24	4.5	5.7	4.3	5.2	3.8	4.7	-21.2	-19.9	NA	NA	18.3	18.9
2021/3/25	4.7	5.7	4.4	5.1	3.9	4.8	-21.2	-20.7	NA	NA	18.2	19.1
2021/3/26	4.7	5.7	4.3	5.2	3.8	4.8	-20.9	-20.3	NA	NA	18.4	18.9
2021/3/27	5.1	5.7	4.4	4.9	4.4	4.8	-20.7	-20.4	NA	NA	18.4	18.6
2021/3/28	5.1	5.7	4.4	4.9	4.4	4.8	-20.7	-20.5	NA	NA	18.3	18.4
2021/3/29	4.7	5.8	4.3	4.9	3.9	4.8	-20.7	-20	NA	NA	18.2	19.3
2021/3/30	4.9	5.7	4.3	5	3.9	4.8	-20.3	-20	NA	NA	18.4	19.2
2021/3/31	4.9	5.7	4.3	4.8	3.9	4.8	-20.3	-19.8	NA	NA	18.6	19.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥 (110)
 師(三)級藥師 040
 藥學部 溫雅如 (110)
 師(三)級藥師 040