

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 03 月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2021/3/1	4.9	5.3	4.9	5.6	19.9	20.2
2021/3/2	4.8	5.3	4.8	5.7	19.9	20.1
2021/3/3	4.8	5.6	4.8	5.7	19.9	20.2
2021/3/4	4.9	5.7	4.8	6.1	20.1	20.4
2021/3/5	5.1	5.6	4.8	6.1	20.4	20.6
2021/3/6	4.9	5.5	4.9	5.6	20.3	20.6
2021/3/7	4.9	5.5	4.9	5.6	20.1	20.3
2021/3/8	4.9	5.7	4.9	5.9	19.9	20.2
2021/3/9	4.7	5.9	4.9	5.8	20.1	20.4
2021/3/10	4.5	5.3	4.9	5.9	20.3	20.5
2021/3/11	4.5	5.3	5.1	6.3	20.4	20.7
2021/3/12	4.6	5.4	5.1	5.8	20.5	20.7
2021/3/13	4.7	5.3	5.1	5.9	20.3	20.7
2021/3/14	4.7	5.3	5.2	5.9	20.1	20.3
2021/3/15	4.7	5.3	5.2	5.9	20.1	20.3
2021/3/16	4.7	5.8	5.3	5.9	20.3	20.6
2021/3/17	4.7	5.3	4.9	6.6	20.4	20.6
2021/3/18	4.7	5.3	4.9	5.8	20.5	20.7
2021/3/19	4.7	5.4	4.9	5.8	20.6	20.8
2021/3/20	4.8	5.4	4.9	5.8	20.6	20.7
2021/3/21	4.8	5.4	4.9	5.8	20.2	20.6
2021/3/22	4.7	5.4	4.8	5.9	19.9	20.2
2021/3/23	4.7	5.5	4.9	5.9	19.8	20.1
2021/3/24	4.7	5.4	5.2	6	20.1	20.3
2021/3/25	4.7	5.3	5.3	6	20.1	20.4
2021/3/26	4.7	5.9	5.2	5.9	20.3	20.6
2021/3/27	4.7	5.3	5.2	5.8	20.3	20.5
2021/3/28	4.7	5.3	5.2	5.9	20.2	20.3
2021/3/29	4.7	5.5	5.1	5.9	20.1	20.4
2021/3/30	4.7	5.7	5.1	6.1	20.4	20.7
2021/3/31	4.7	5.6	5.4	6.4	20.6	20.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥  
師(三)級藥師 1100401 藥學部 四雅如  
師(三)級藥師 1100406

溫度均在正常範圍內