

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 04 月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/4/1	4.8	5.7	4.2	4.8	3.9	4.9	-19.6	-19.4	NA	NA	18.6	19.2
2021/4/2	5.1	5.7	4.3	4.8	4.4	4.9	-19.6	-19.6	NA	NA	18.6	18.8
2021/4/3	5.1	5.7	4.3	4.8	4.5	4.9	-19.7	-19.6	NA	NA	18.4	18.6
2021/4/4	5.1	5.7	4.3	4.8	4.4	4.9	-19.7	-19.6	NA	NA	18.3	18.4
2021/4/5	5.1	5.7	4.3	4.8	4.4	4.9	-19.7	-19.6	NA	NA	18.1	18.6
2021/4/6	4.9	5.6	4.3	4.9	3.9	4.9	-19.9	-18.9	NA	NA	18.6	19.4
2021/4/7	4.8	5.7	4.3	5.1	3.9	4.9	-20.2	-19.6	NA	NA	18.8	19.6
2021/4/8	4.8	5.7	4.3	4.9	3.9	4.9	-20.2	-19.4	NA	NA	19	19.7
2021/4/9	4.8	5.7	4.3	4.9	3.8	4.8	-20.1	-19.6	NA	NA	19.1	19.7
2021/4/10	5.1	5.7	4.3	4.8	4.3	4.7	-20.1	-19.8	NA	NA	19	19.2
2021/4/11	5.1	5.7	4.3	4.8	4.4	4.7	-20.1	-19.9	NA	NA	18.9	19
2021/4/12	4.8	5.8	4.3	4.8	3.7	4.7	-21.2	-19.9	NA	NA	18.8	19.5
2021/4/13	4.8	5.7	4.4	5.1	3.8	4.7	-21.2	-20.4	NA	NA	19	19.6
2021/4/14	4.9	5.7	4.3	4.9	3.8	4.8	-21.2	-20.5	NA	NA	19.1	19.8
2021/4/15	4.2	5.8	4.3	4.9	3.8	4.9	-20.9	-20.5	NA	NA	19.2	19.8
2021/4/16	4.2	5.3	4.3	4.8	3.9	4.9	-20.9	-20.6	NA	NA	19.2	19.8
2021/4/17	4.2	4.8	4.3	4.8	4.4	4.8	-20.9	-20.7	NA	NA	19.1	19.4
2021/4/18	4.2	4.8	4.3	4.8	4.4	4.8	-20.9	-20.7	NA	NA	18.9	19.2
2021/4/19	4.2	5.3	4.4	4.9	4.1	4.7	-20.9	-20.2	NA	NA	18.9	19.5
2021/4/20	4.2	5.2	4.4	5.1	3.9	4.7	-20.5	-19.6	NA	NA	19	19.7
2021/4/21	4.2	5.1	4.3	5.2	3.9	4.9	-20.7	-19.9	NA	NA	19.2	19.8
2021/4/22	4.2	5.2	4.2	4.9	4.1	5.1	-20.7	-19.5	NA	NA	19.2	19.9
2021/4/23	3.7	5.9	4.3	5.1	4.1	4.9	-20.4	-19.9	NA	NA	19.3	19.9
2021/4/24	4.1	5.1	4.4	4.9	4.6	4.9	-20.4	-20.2	NA	NA	19.2	19.4
2021/4/25	4.1	5	4.4	4.9	4.6	4.9	-20.4	-20.2	NA	NA	19.1	19.2
2021/4/26	4.1	5.1	4.2	5.1	4.2	4.9	-20.6	-19.1	NA	NA	19	19.6
2021/4/27	3.9	5.9	4.3	5.1	4.2	4.9	-19.7	-19.1	NA	NA	19.2	19.9
2021/4/28	3.3	5.7	4.3	5.1	4.2	5	-20.2	-19.6	NA	NA	19.2	19.8
2021/4/29	3.4	5.6	4.3	5.2	4.1	4.9	-20.2	-19.4	NA	NA	19.3	21.8
2021/4/30	3.6	4.9	4.2	5.1	4.1	4.8	-20.6	-19.6	NA	NA	19.2	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥 (110)
師(三)級藥師 (0503)
藥學部 溫雅如 (110)
師(三)級藥師 (0503)