

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 04月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8℃		2~8℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max
2021/4/1	4.4	5	5.3	6.3	20.7	21.1
2021/4/2	4.4	5	5.4	6	20.7	20.9
2021/4/3	4.4	5	5.3	6	20.6	20.7
2021/4/4	4.5	5	5.4	6	20.3	20.6
2021/4/5	4.5	5.1	5.3	6	20.1	20.3
2021/4/6	4.4	5.1	5.4	6.3	19.9	20.4
2021/4/7	4.4	5.3	5.3	6.2	20.3	20.7
2021/4/8	4.7	5.3	5.3	6.1	20.6	21.1
2021/4/9	4.7	5.4	5.3	6.2	20.8	21.1
2021/4/10	4.8	5.3	5.3	5.9	20.7	20.9
2021/4/11	4.8	5.3	5.3	5.9	20.6	20.8
2021/4/12	4.5	5.3	5.2	5.9	20.5	20.8
2021/4/13	4.4	5.2	5.2	5.9	20.7	21.1
2021/4/14	4.4	5.8	5.2	5.9	20.9	21.2
2021/4/15	4.4	5.2	5.3	5.9	21.1	21.3
2021/4/16	4.4	5.2	5.3	5.9	21.1	21.3
2021/4/17	4.4	5.1	5.3	5.9	21.1	21.3
2021/4/18	4.4	5.1	5.3	5.9	20.6	21.1
2021/4/19	4.4	5.2	5.3	6	20.5	20.9
2021/4/20	4.4	5.3	5.3	6.1	20.7	21.1
2021/4/21	4.4	5.3	5.3	6	20.9	21.3
2021/4/22	4.5	5.2	5.3	5.9	21.1	21.4
2021/4/23	4.4	5.3	5.3	6.2	21.3	21.6
2021/4/24	4.5	5.2	5.7	6.1	21.3	21.5
2021/4/25	4.5	5.2	5.7	6.1	21.1	21.3
2021/4/26	4.5	5.7	5.1	6.1	20.9	21.2
2021/4/27	4.6	5.6	5.2	6.7	21.1	21.4
2021/4/28	4.8	5.5	6	7	21.2	21.4
2021/4/29	4.8	5.5	5.7	6.6	21.3	21.7
2021/4/30	4.9	5.5	5.2	6.3	21.3	21.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 110 2803 1416  
 藥學部 溫雅如 110 0503 1641