

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年05月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/5/1	3.7	4.9	4.4	4.9	4.4	4.8	-20.4	-20.2	NA	NA	19.2	19.4
2021/5/2	3.7	4.8	4.4	4.9	4.4	4.9	-20.4	-20.2	NA	NA	19.1	19.2
2021/5/3	3.1	4.8	4.2	5.1	4	4.8	-19.1	-18.8	NA	NA	19.2	19.6
2021/5/4	3.5	4.9	4.3	5	4	4.8	-20.3	-18.9	NA	NA	19.2	19.7
2021/5/5	3.8	5.1	4.3	4.9	4.1	4.8	-20.4	-19.9	NA	NA	19.2	19.8
2021/5/6	3.7	5	4.2	4.8	4	4.9	-20.4	-17.9	NA	NA	19.2	19.8
2021/5/7	3.7	4.9	4.2	4.8	4.1	4.9	-20.4	-19.3	NA	NA	19.2	19.8
2021/5/8	3.7	4.9	4.4	4.8	4.4	4.9	-20.4	-20.2	NA	NA	19.3	19.4
2021/5/9	3.7	4.9	4.3	4.8	4.4	4.9	-20.4	-20.2	NA	NA	19.2	19.3
2021/5/10	3.7	4.9	4.1	4.8	4.1	4.9	-20.4	-19.6	NA	NA	19.2	19.7
2021/5/11	3.8	5.3	3.9	5.1	4.1	4.9	-20.2	-18.8	NA	NA	19.2	19.8
2021/5/12	3.7	4.8	3.9	4.6	4.1	6.4	-19.8	-18.8	NA	NA	19.3	19.9
2021/5/13	3.7	4.7	3.9	4.5	4.3	4.9	-19.8	-19.1	NA	NA	19.4	19.9
2021/5/14	3.7	4.8	4.1	4.8	4.3	4.9	-19.6	-18.8	NA	NA	19.3	19.8
2021/5/15	3.7	4.8	4.3	4.8	4.4	4.7	-19.4	-19.3	NA	NA	19.3	19.4
2021/5/16	3.7	4.8	4.3	4.8	4.4	4.7	-19.4	-19.3	NA	NA	19.2	19.3
2021/5/17	3.1	4.7	4.3	4.9	4.4	4.8	-19.4	-18.7	NA	NA	19.2	19.7
2021/5/18	3.3	4.7	4.3	4.8	4.2	4.8	-20.7	-19.1	NA	NA	19.2	19.8
2021/5/19	3.7	4.8	4.2	4.8	4.2	4.7	-20.7	-19.9	NA	NA	19.3	19.8
2021/5/20	3.7	5.3	4.2	4.8	4.3	4.9	-20.7	-19.1	NA	NA	19.3	19.8
2021/5/21	3.6	4.8	4	5.1	4.3	4.8	-20.7	-20.3	NA	NA	19.2	19.7
2021/5/22	3.6	4.8	4	4.5	4.4	4.8	-20.7	-20.4	NA	NA	19.1	19.3
2021/5/23	3.6	4.8	4	4.5	4.4	4.8	-20.7	-20.4	NA	NA	19	19.1
2021/5/24	3.6	4.8	4	4.5	4.3	4.8	-20.7	-19.6	NA	NA	18.8	19.4
2021/5/25	3.6	4.8	3.9	4.6	4.3	4.8	-19.9	-17.7	NA	NA	19	19.6
2021/5/26	3.7	4.8	4	4.8	4.4	4.9	-19.8	-19	NA	NA	18.8	19.6
2021/5/27	3.3	4.9	4	4.6	4.4	4.9	-20.3	-19.6	NA	NA	19	19.6
2021/5/28	3.4	4.8	4	4.7	4.2	4.9	-20.4	-19.1	NA	NA	19.1	19.6
2021/5/29	3.4	4.7	4.3	4.7	4.4	4.8	-19.6	-19.6	NA	NA	19.1	19.2
2021/5/30	3.4	4.7	4.2	4.7	4.4	4.8	-20.7	-18.1	NA	NA	18.9	19
2021/5/31	3.3	4.6	4.2	4.7	4.2	4.8	-20.7	-20.4	NA	NA	18.9	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 1100601  
藥學部 溫雅如 1100601

溫度均在正常範圍內