

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 05月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2021/5/1	4.9	5.5	5.3	5.9	21.1	21.4
2021/5/2	4.9	5.4	5.3	5.9	20.8	21.1
2021/5/3	4.9	5.5	5.7	6.2	20.7	21.1
2021/5/4	4.9	5.4	5.7	6.3	20.9	21.3
2021/5/5	4.9	5.7	5.6	6.3	21.1	21.5
2021/5/6	4.9	5.5	5.3	6.2	21.3	21.7
2021/5/7	4.9	5.5	5.4	6.2	21.4	21.7
2021/5/8	4.9	5.4	5.7	6.2	21.3	21.6
2021/5/9	4.9	5.4	5.7	6.2	21.1	21.3
2021/5/10	4.9	5.6	5.6	6.3	20.9	21.3
2021/5/11	4.9	5.4	5.2	6.2	21.2	21.4
2021/5/12	4.9	5.7	5.2	5.9	21.3	21.7
2021/5/13	4.9	5.4	5.2	5.8	21.6	21.8
2021/5/14	4.6	6	5.2	5.8	21.6	21.8
2021/5/15	4.7	5.2	5.2	5.7	21.5	21.7
2021/5/16	4.6	5.1	5.2	5.7	21.2	21.5
2021/5/17	4.4	5.1	5.2	5.8	21.1	21.4
2021/5/18	4.4	5.1	5.1	6	21.3	21.4
2021/5/19	4.4	5.2	4.9	6	21.3	21.6
2021/5/20	4.6	5.1	4.9	5.9	21.3	21.6
2021/5/21	4.6	5.1	4.9	6	21.3	21.7
2021/5/22	4.6	5.1	5.1	5.7	21.3	21.6
2021/5/23	4.6	5.1	5.1	5.7	20.9	21.3
2021/5/24	4.6	5.2	5.1	6.2	20.8	21.1
2021/5/25	4.6	5.3	5.1	6.1	20.9	21.2
2021/5/26	4.7	5.3	5.1	6.2	21.1	21.2
2021/5/27	4.8	5.3	5.2	6	21.1	21.3
2021/5/28	4.7	5.4	5.2	6	21.1	21.3
2021/5/29	4.7	5.3	5.2	6	21.1	21.3
2021/5/30	4.8	5.3	5.2	6	20.8	21.1
2021/5/31	4.7	5.3	5.2	6	20.7	21.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 (三)級藥師 1100601

藥學部 溫雅如 (三)級藥師 1100601