

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 06 月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362												
0850-0835												
2~8°C												
2~8°C												
2~8°C												
負15~負25°C												
≤負70°C												
15~25°C												
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/6/1	3.3	4.6	4	4.7	4.4	4.9	-20.9	-18.9	NA	NA	19.7	20.1
2021/6/2	3.1	4.7	4.1	4.8	4.4	5.1	-20.7	-18.9	NA	NA	19.3	19.9
2021/6/3	3.1	4.8	4	4.5	4.4	5.1	-20.1	-19.5	NA	NA	19.2	19.7
2021/6/4	3.4	4.8	4	4.7	4.4	5.9	-20.4	-19.8	NA	NA	19.2	19.8
2021/6/5	3.4	4.8	4.2	4.6	4.7	5.1	-20.4	-20.1	NA	NA	19.2	19.4
2021/6/6	3.4	4.8	4.2	4.6	4.8	5.1	-20.4	-20.1	NA	NA	19	19.3
2021/6/7	3.4	5.1	4.2	4.7	4.4	5.2	-20.4	-19.6	NA	NA	19.3	19.9
2021/6/8	3.5	4.8	4.2	4.7	4.4	5.2	-20.4	-19.9	NA	NA	19.2	19.6
2021/6/9	3.4	4.8	4	4.6	4.4	5.2	-20.7	-19.9	NA	NA	19.2	19.6
2021/6/10	3.3	4.9	4.1	4.6	4.5	5.2	-20.8	-19.9	NA	NA	19.2	19.6
2021/6/11	3.3	4.6	4	4.7	4.3	5.3	-20.9	-20	NA	NA	19.2	19.7
2021/6/12	3.3	4.6	4.1	4.5	4.9	5.2	-20.9	-20.4	NA	NA	19.2	19.9
2021/6/13	3.4	4.7	4.1	4.5	4.9	5.2	-20.9	-20.5	NA	NA	18.6	19.3
2021/6/14	3.4	4.7	4.1	4.4	4.9	5.2	-20.9	-20.4	NA	NA	18.4	18.6
2021/6/15	3.4	4.7	4.1	4.7	4.5	5.3	-20.9	-19.5	NA	NA	18.4	20.1
2021/6/16	3.6	4.8	4.1	4.6	4.4	5.3	-20.4	-19.9	NA	NA	19.7	20.3
2021/6/17	3.6	4.8	4.2	4.7	4.3	5.3	-20.6	-20.2	NA	NA	19.8	20.4
2021/6/18	3.6	4.9	4.1	4.7	4.3	5.1	-20.6	-19.8	NA	NA	19.4	20.4
2021/6/19	3.6	4.9	4.1	4.5	4.6	4.9	-20.2	-19.9	NA	NA	18.8	19.4
2021/6/20	3.7	4.9	4.1	4.5	4.6	4.9	-20.2	-19.9	NA	NA	18.6	18.8
2021/6/21	3.7	5	4.1	4.9	4.4	5.1	-20.2	-19.4	NA	NA	18.4	19.4
2021/6/22	3.4	5.1	4.1	4.9	4.4	5.1	-20.1	-19.6	NA	NA	18.6	19.1
2021/6/23	3.8	5.1	4.1	4.9	4.4	5.1	-19.8	-19.3	NA	NA	18.9	20.8
2021/6/24	3.7	5.1	4.1	4.7	4.4	5.2	-20.4	-19.6	NA	NA	19.2	19.7
2021/6/25	3.8	5.1	4	4.7	4.4	5.1	-20.7	-20.2	NA	NA	19.2	19.8
2021/6/26	3.9	5	4	4.4	4.7	5.1	-20.9	-20.2	NA	NA	18.8	19.5
2021/6/27	3.9	5.1	4	4.4	4.7	5.1	-20.9	-20.4	NA	NA	18.6	18.8
2021/6/28	3.9	5.1	3.9	4.7	3.8	6.1	-21	-20.3	NA	NA	18.4	19.1
2021/6/29	3.8	5	3.9	4.4	2.9	4.4	-21.4	-20.7	NA	NA	18.6	19.1
2021/6/30	3.9	5.1	3.8	4.5	3.6	4.9	-21.4	-20.8	NA	NA	18.1	18.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 110
師(三)級藥師 0101
藥學部 邱子芹 110
師(三)級藥師 0101