

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年07月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362	0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007						
0850-0835	0850-0835	0520-293	1504-0339	1213-0010	1330-0064	1405-0194						
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2021/7/1	4	4.9	4.2	5	4.6	5.3	-20.7	-19	NA	NA	18.2	22.8
2021/7/2	3.9	5	4.3	4.7	4.5	5.2	-20.7	-20.1	NA	NA	18	18.6
2021/7/3	3.9	4.8	4.4	4.7	4.5	4.8	-20.6	-20.4	NA	NA	17.9	18.2
2021/7/4	3.9	4.9	4.5	4.7	4.5	4.8	-20.6	-20.3	NA	NA	17.8	17.9
2021/7/5	3.9	4.9	4.3	4.9	4.4	5.3	-20.7	-20.2	NA	NA	17.7	18.6
2021/7/6	4	5	4.4	4.9	4.4	5.1	-20.7	-19.8	NA	NA	17.9	19
2021/7/7	4.2	5.1	4.4	4.8	4.4	5.3	-20.4	-19.9	NA	NA	18.2	18.9
2021/7/8	4	5.1	4.4	5.1	4.3	5.6	-20.4	-19.9	NA	NA	18.2	18.9
2021/7/9	4.1	4.8	4.4	5.1	4.3	5.2	-20.7	-20.3	NA	NA	18.3	18.9
2021/7/10	4.2	4.8	4.6	4.8	4.3	4.7	-20.7	-20.4	NA	NA	18.2	18.4
2021/7/11	4.2	4.9	4.6	4.8	4.3	4.7	-20.7	-20.4	NA	NA	18.2	18.2
2021/7/12	4	5.1	4.3	4.8	4.3	5.3	-20.7	-19.9	NA	NA	18.1	19.1
2021/7/13	4	5.3	4.3	5.3	4.3	5.2	-20.7	-20.3	NA	NA	18.2	19.7
2021/7/14	4.2	5.1	4	4.8	4.3	5.3	-20.7	-20.1	NA	NA	18.3	19.2
2021/7/15	4.2	5.1	4	4.5	4.4	5.4	-20.6	-20.2	NA	NA	18.4	19.2
2021/7/16	3.9	5.1	4	4.4	4.4	5.8	-20.7	-20	NA	NA	18.4	19.1
2021/7/17	4.2	4.9	4.1	4.4	4.4	4.7	-20.3	-20	NA	NA	18.4	18.6
2021/7/18	4.2	4.9	4.1	4.4	4.4	4.7	-20.3	-20.1	NA	NA	18.3	18.4
2021/7/19	4.2	5.1	4	4.5	4.4	5.6	-20.3	-20.1	NA	NA	18.2	19.3
2021/7/20	4.2	5.1	4.1	5.1	4.4	5.5	-20.3	-19.9	NA	NA	19.1	19.9
2021/7/21	4.3	5.3	4.1	4.5	4.4	5.5	-20.1	-19.4	NA	NA	19.1	19.9
2021/7/22	4.3	5.1	4.2	5.1	4.4	5.6	-20.2	-19.7	NA	NA	19.1	19.9
2021/7/23	3.6	5.6	4.4	5	4.4	5.6	-20.2	-19.8	NA	NA	19.2	19.8
2021/7/24	4.2	4.8	4.4	4.7	4.4	4.7	-20.3	-20	NA	NA	19.2	19.4
2021/7/25	4.2	4.8	4.4	4.7	4.4	4.7	-20.3	-20.1	NA	NA	19.1	19.2
2021/7/26	3.9	4.9	4.4	4.8	4.3	5.6	-20.3	-19.8	NA	NA	19	20.3
2021/7/27	3.2	5.1	4.2	4.8	4.3	5.2	-20.2	-19.7	NA	NA	19.1	20
2021/7/28	3.4	4.6	4.2	4.7	4.3	5.3	-20.1	-19.6	NA	NA	19.2	20.1
2021/7/29	3.4	5.1	4.2	4.6	4.3	5.2	-20.1	-19.7	NA	NA	19.3	20.1
2021/7/30	3.5	4.6	4.2	4.6	4.4	5.1	-20.1	-19.6	NA	NA	19.4	20.2
2021/7/31	3.7	4.7	4.2	4.5	4.3	4.8	-19.9	-19.8	NA	NA	19.3	19.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 110 2802
藥學部 邱子芹 110 2802