

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110 年 07 月

	四號冰箱		五號冰箱		室溫2	
序號	1611-0062		1715-0118		1715-0116	
備用	1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2021/7/1	4.8	5.4	4.4	5.3	20.5	20.7
2021/7/2	4.9	5.4	4.4	5.5	20.5	20.7
2021/7/3	4.9	5.4	4.5	5.4	20.6	20.7
2021/7/4	4.9	5.4	4.6	5.5	20.3	20.6
2021/7/5	4.8	5.7	4.5	5.7	20.2	20.6
2021/7/6	4.7	5.6	4.6	5.9	20.3	20.7
2021/7/7	4.7	6.1	4.2	5.8	20.6	21.1
2021/7/8	4.7	5.4	4.6	5.7	20.9	21.3
2021/7/9	4.7	5.7	4.8	5.7	21.1	21.3
2021/7/10	4.9	5.4	4.8	5.7	21.2	21.3
2021/7/11	4.9	5.4	4.8	5.7	20.9	21.1
2021/7/12	4.7	5.4	4.8	5.9	20.8	21.1
2021/7/13	4.7	6.4	5.1	6.1	21.1	21.3
2021/7/14	4.7	5.7	4.8	5.9	21.2	21.4
2021/7/15	4.8	5.4	5.2	6.9	21.3	21.6
2021/7/16	4.9	5.8	4.7	6.1	21.3	21.5
2021/7/17	5.3	5.9	5.3	6.1	21.3	21.5
2021/7/18	5.4	5.8	5.3	6.1	21.2	21.3
2021/7/19	4.7	5.8	4.5	6.1	21.1	21.4
2021/7/20	4.5	5.3	5.2	6.1	21.3	21.7
2021/7/21	4.5	5.7	4.9	6.2	21.7	21.9
2021/7/22	4.9	6.1	5.2	6.1	21.8	22.1
2021/7/23	5.1	5.9	5	7.2	22	22.2
2021/7/24	5.2	5.7	4.9	5.7	22	22.2
2021/7/25	5.1	5.7	5	5.7	21.7	22
2021/7/26	5.1	5.7	4.5	5.6	21.7	21.9
2021/7/27	5.2	5.8	4.4	5.6	21.7	22
2021/7/28	4.2	5.8	4.7	5.7	21.9	22.3
2021/7/29	4.3	5.4	4.8	5.7	22.1	22.3
2021/7/30	4.7	5.5	4.8	5.7	22.2	22.4
2021/7/31	4.7	5.2	4.9	5.7	22.2	22.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 110 2802
師(三)級藥師 藥學部 邱子芹 110 2802
師(三)級藥師