

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 110年 08 月

	零號冰箱		四號冰箱		五號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		1715-0116	
備用	2052-0129		1611-0052		1715-0119		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max
2021/8/1	6.2	6.6	4.7	5.2	4.9	5.7	21.9	22.2
2021/8/2	6.1	6.6	4.7	5.2	5	7.1	21.8	22.1
2021/8/3	5.9	6.4	4.7	5.3	4.9	5.9	21.9	22.2
2021/8/4	5.9	6.5	4.7	5.3	4.9	5.8	21.9	22.5
2021/8/5	6	6.4	4.7	5.3	4.7	5.7	22.1	22.4
2021/8/6	6.1	6.5	4.8	5.3	4.7	5.7	21.9	22.2
2021/8/7	6.1	6.4	4.8	5.3	4.9	5.7	21.7	22.1
2021/8/8	6	6.3	4.8	5.2	4.9	5.7	21.3	21.7
2021/8/9	5.9	6.7	4.8	5.3	4.9	5.8	21.3	21.7
2021/8/10	5.9	6.9	4.8	5.3	4.9	6.1	21.6	21.9
2021/8/11	5.4	6.3	4.8	5.7	4.9	5.9	21.7	22
2021/8/12	5.3	6.1	5.3	5.7	5.1	6	21.8	22.2
2021/8/13	5.2	7.2	5.3	5.8	4.7	5.9	22.1	22.2
2021/8/14	5.4	6.1	5.3	5.7	4.7	5.7	21.8	22.1
2021/8/15	5.4	6	5.2	5.7	4.7	5.7	21.4	21.8
2021/8/16	5.1	5.9	5.2	5.7	4.7	5.7	21.3	21.7
2021/8/17	5.1	5.9	4.7	5.7	4.7	6.2	21.5	21.9
2021/8/18	5.2	6.4	4.7	5.3	4.7	5.7	21.7	22
2021/8/19	5.2	6.1	4.7	5.6	4.8	5.7	21.8	22
2021/8/20	5.3	5.9	4.6	5.3	4.9	5.8	21.9	22
2021/8/21	5.4	5.9	4.5	4.9	5.1	5.8	21.7	21.9
2021/8/22	5.4	5.9	4.5	4.9	5.1	5.8	21.4	21.7
2021/8/23	5.4	5.9	4.5	4.9	5.2	5.9	21.3	21.7
2021/8/24	5.3	5.9	4.5	5.4	5.3	5.9	21.5	21.9
2021/8/25	5.3	6.1	4.5	5.5	5.3	6	21.7	22
2021/8/26	5.1	6.1	4.9	5.6	5.2	5.9	21.8	22.2
2021/8/27	5.1	5.8	5.2	5.6	5.3	5.9	21.9	22.4
2021/8/28	5.1	5.7	5.2	5.5	5.3	5.9	21.9	22.2
2021/8/29	5.1	5.7	5.1	5.4	5.3	5.9	21.7	21.9
2021/8/30	5.1	5.9	4.8	5.4	5.3	5.9	21.5	21.9
2021/8/31	4.9	5.7	4.8	5.3	5.2	5.9	21.8	22.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 110/8/31
藥學部 邱子芹 110/8/31

溫度均在正常範圍內