

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 110 年 08 月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2021/8/1	4.7	5.2	20.3	20.9
2021/8/2	4.6	5.3	20.2	20.8
2021/8/3	4.6	5.2	20.5	21.2
2021/8/4	4.6	5.3	20.4	21.1
2021/8/5	4.7	5.2	20.6	21.1
2021/8/6	4.7	5.2	20.4	21.1
2021/8/7	4.7	5.2	20.5	21
2021/8/8	4.7	5.2	20.4	20.8
2021/8/9	4.7	5.2	20.5	21.2
2021/8/10	4.7	5.3	20.3	21.1
2021/8/11	4.7	5.3	20.3	21.3
2021/8/12	4.7	5.2	20.6	21.3
2021/8/13	4.7	5.3	20.7	21.3
2021/8/14	4.9	5.3	20.6	20.8
2021/8/15	4.9	5.3	20.3	21
2021/8/16	4.8	5.4	20.2	21.1
2021/8/17	4.9	5.4	20.3	21.2
2021/8/18	4.7	5.6	20.4	21.1
2021/8/19	4.7	5.5	20.6	21.2
2021/8/20	4.7	5.3	20.8	21.2
2021/8/21	4.7	5.2	20.5	21.1
2021/8/22	4.7	5.3	20.3	20.6
2021/8/23	4.7	5.3	20.1	21.2
2021/8/24	4.7	5.3	20.4	21.2
2021/8/25	4.7	5.2	20.5	21.4
2021/8/26	4.7	5.2	20.6	21.1
2021/8/27	4.7	5.3	20.5	21.2
2021/8/28	4.8	5.3	20.8	21.1
2021/8/29	4.8	5.3	20.6	20.9
2021/8/30	4.7	5.3	20.6	21.2
2021/8/31	4.6	5.2	20.9	21.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 110
師(三級藥師) 邱保祥 0901
藥學部 邱予芹 110
師(三級藥師) 邱予芹 0901